

AS IV Serial no: 111778
Version no: 532B

TEST RECORD 00502

Temp Date Time 210L

Air Blank:
10/05/23 23:33 .000
Calibration Check:
21 10/05/23 23:33 .101

Subject Name

Test 1

Subject I.D.

Bryan 2202

Operator Name, I.D.
230159 8-7-25

Location

2110 S. Broadway

Oak Grove MO
64075

AS IV Serial no: 111778
Version no: 532B

TEST RECORD 00503

Temp Date Time 210L

Air Blank:
10/05/23 23:35 .000
Calibration Check:
22 10/05/23 23:35 .100

Subject Name

Test 2

Subject I.D.

Bryan 2202

Operator Name, I.D.
230159 8-7-25

Location

2110 S. Broadway

Oak Grove MO
64075

AS IV Serial no: 111778
Version no: 532B

TEST RECORD 00504

Temp Date Time 210L

Air Blank:
10/05/23 23:36 .000
Calibration Check:
23 10/05/23 23:36 .100

Subject Name

Test 3

Subject I.D.

Bryan 2202

Operator Name, I.D.
230159 8-7-25

Location

2110 S. Broadway

Oak Grove MO
64075

AS IV Serial no: 111778
Version no: 532B

TEST RECORD 00505

Temp Date Time 210L

VOID: RFI
12 10/05/23 23:38

Subject Name

RFI Test

Subject I.D.

Bryan 2202

Operator Name, I.D.
230159 8-7-25

Location

2110 S. Broadway

Oak Grove MO
64075



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22310** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **August 16, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1205%** (w/vol) ethyl alcohol. The expiration date for this lot number is **August 11, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RANDALL BRYAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/7/2023

Mike Mussum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230159

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/7/2025

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BRYAN, RANDALL
 Permit No 230159
 Date issued 8/7/2023 Date Expires 8/7/2025

