



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Dwyer at 7:56 am, Jun 12, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111779	PRINTER SN 09B.3589.547	DATE OF INSPECTION 06/07/2023
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 110 South Maple St Galena	TIME OF INSPECTION 3:44 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth LOT # 22430 EXP. DATE 11/30/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP6027 SIMULATOR EXP DATE 02/27/2024

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .105

TEST 2 .104

TEST 3 .104

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Blank test prior to maintenance

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Kellyn Bailey
---------------	-----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 230099 5/30/2025	TELEPHONE NUMBER (417) 357-6116
---	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

B

AS IV Serial no: 111779
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00660

Temp Date Time ^{s/} 210L
Air Blank:
06/07/23 15:44 .000
Subject Test: Auto
21 06/07/23 15:44 .000

Subject Name
Self test
Subject I.D.
Monthly Maint
Operator Name, I.D.
K. Bailey
Location
Stare Co Jail

AS IV Serial no: 111779
Version no: 532B

TEST RECORD 00661

Temp Date Time ^{s/} 210L
Air Blank:
06/07/23 15:48 .000
Calibration Check:
22 06/07/23 15:48 .185

Subject Name
Test 4
Subject I.D.
Monthly Maint
Operator Name, I.D.
K. Bailey
Location
Stare Co Jail

AS IV Serial no: 111779
Version no: 532B

TEST RECORD 00662

Temp Date Time ^{s/} 210L
Air Blank:
06/07/23 15:50 .000
Calibration Check:
23 06/07/23 15:50 .104

Subject Name
Test 2
Subject I.D.
Monthly Maint
Operator Name, I.D.
K. Bailey
Location
Stare Co Jail

AS IV Serial no: 111779
Version no: 532B

TEST RECORD 00663

Temp Date Time ^{s/} 210L
Air Blank:
06/07/23 15:53 .000
Calibration Check:
23 06/07/23 15:53 .164

Subject Name
Test 3
Subject I.D.
Monthly Maint
Operator Name, I.D.
K. Bailey
Location
Stare Co Jail

AS IV Serial no: 111779
Version no: 532B

TEST RECORD 00664

Temp Date Time ^{s/} 210L
VOID: RFI
12 06/07/23 15:55

Subject Name
RFI
Subject I.D.
Monthly Maint
Operator Name, I.D.
K. Bailey
Location
Stare Co Jail



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
KELLYN BAILEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/30/2023

NUMBER 230099

EXPIRES 5/30/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave I. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BAILEY, KELLYN
 Permit No 230099
 Date Issued 5/30/2023 Date Expires 5/30/2025

