



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|--|----------------------------------|
| ALCO SENSOR IV SN 111780 | NAME OF AGENCY KANSAS CITY MISSOURI POLICE DEPT | DATE OF INSPECTION 01/01/2023 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 9701 MARION PARK DRIVE, KANSAS CITY, MO 64137 | | TIME OF INSPECTION 2120 |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u> | LOT # <u>AG132803</u> EXP. DATE <u>11/24/2023</u> |
| <input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) | SIM. SN _____ SIM. NIST EXP DATE _____ |

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|--------------|--------------|--------------|
| TEST 1 .080 | TEST 2 .080 | TEST 3 .080 |
|--------------|--------------|--------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|----------|--------------|----------------|-----------|----------------|------------|
| REFUSALS | (0-.04) 1 | (.05-.09) 2 | (.10-.14) | (.15-.19) 1 | (OVER .19) |
|----------|--------------|----------------|-----------|----------------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT MEETS ALL DHSS STANDARDS AND GUIDELINES.

INSPECTING OFFICER

| | |
|---|---|
| SIGNATURE #5844 | PRINT NAME P.O. MATTHEW FRANCOIS #5844 |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220275 -12/21/2024 | TELEPHONE NUMBER (816) 766-5943 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00729

| Temp | Date | Time | g/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
01/01/23 21:27 .000
Calibration Check:
20 01/01/23 21:27 .080

Subject Name
TEST

Subject I.D.
1

Operator Name, I.D.
FRANCOIS 220275

Location
9701 MARION PARK DR

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00730

| Temp | Date | Time | g/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
01/01/23 21:29 .000
Calibration Check:
22 01/01/23 21:29 .080

Subject Name
TEST

Subject I.D.
2

Operator Name, I.D.
FRANCOIS 220275

Location
9701 MARION PARK DR

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00731

| Temp | Date | Time | g/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
01/01/23 21:30 .000
Calibration Check:
23 01/01/23 21:30 .080

Subject Name
TEST

Subject I.D.
3

Operator Name, I.D.
FRANCOIS 220275

Location
9701 MARION PARK DR

AS IV Serial no: 111780
Version no: 532B

TEST RECORD 00732

| Temp | Date | Time | g/ 210L |
|------|------|------|------------|
|------|------|------|------------|

VOID: RFI
12 01/01/23 21:32

Subject Name
TEST

Subject I.D.
4

Operator Name, I.D.
FRANCOIS 220275

Location
9701 MARION PARK DR



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 29-Nov-2021

Lot # AG132803 **Model** 108

| | | | |
|-----------------|------------------|---------------------|--------------------------------|
| Exp Date | Cyl. Type | Component | Certified Concentration |
| 24-Nov-2023 | 108 | Ethanol Nitrogen | 0.082 ± 0.002 BrAC (223 ppm) |

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581 | 392.1 ppm | EB0010603 | 393.0 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.2 ppm |
| EB0010285 | 208.0 ppm | EB0010595 | 208.3 ppm |
| EB0010561 | 103.6 ppm | EB0010562 | 104.2 ppm |
| EB0010681 | 52.12 ppm | EB0010579 | 52.81 ppm |

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC434668 | 800.0 ppm | 0056649 | 390.1 ppm |
| CC234503 | 253.0 ppm | 0056662 | 150.2 ppm |

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:11.30.2021 13:42

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MATTHEW FRANCOIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2022

Mike Massman
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220275

Paula J. Nickelson
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 12/21/2024

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FRANCOIS, MATTHEW
Permit No 220275
Date Issued 12/21/2022 **Date Expires** 12/21/2024

