



RECEIVED

By Tracy Crews at 12:45 pm, Nov 06, 2023

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 119 727	NAME OF AGENCY Garden City Police	DATE OF INSPECTION 11/06/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 106 W 3rd St	Garden City MO 64747	TIME OF INSPECTION 1110 AM

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **Guth Laboratories** LOT # **23180** EXP. DATE **05/17/2025**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34°C** SIM. SN **MP6021** SIM. NIST EXP DATE **03/07/24**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .096	TEST 2 = .096	TEST 3 = .097
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE Mishon Clark	PRINT NAME Mishon Clark
TYPE II PERMIT NUMBER/EXPIRATION DATE 230160 08/07/2025	TELEPHONE NUMBER 816 773 8201

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119727
Version no: 532C

TEST RECORD 00016

Temp Date Time 9/
210L

Air Blank:
11/06/23 11:10 .000
Calibration Check:
24 11/06/23 11:10 .096

Subject Name

Garden City Police

Subject I.D.

Test

Operator Name, I.D.

Mishon Clark ~~230160~~

Location 230160

106 N 3rd

Garden City MO 64747

AS IV Serial no: 119727
Version no: 532C

TEST RECORD 00017

Temp Date Time 9/
210L

Air Blank:
11/06/23 11:14 .000
Calibration Check:
25 11/06/23 11:14 .096

Subject Name

Garden City Police

Subject I.D.

Test

Operator Name, I.D. 230160

Mishon Clark ~~230160~~

Location

106 N 3rd

Garden City MO 64747

AS IV Serial no: 119727
Version no: 532C

TEST RECORD 00018

Temp Date Time 9/
210L

Air Blank:
11/06/23 11:15 .000
Calibration Check:
26 11/06/23 11:15 .097

Subject Name

Garden City Police

Subject I.D.

Test

Operator Name, I.D.

Mishon Clark ~~230160~~

Location 230160

106 N 3rd

Garden City MO 64747

AS IV Serial no: 119727
Version no: 532C

TEST RECORD 00019

Temp Date Time 9/
210L

VOID: RFI
12 11/06/23 11:17

Subject Name

Garden City Police

Subject I.D.

Test

Operator Name, I.D. 230160

Mishon Clark ~~230160~~

Location

106 N 3rd

Garden City MO 64747



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MISHON CLARK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/7/2023

NUMBER 230160

EXPIRES 8/7/2025

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CLARK, MISHON
Permit No 230160
Date Issued 8/7/2023 **Date Expires** 8/7/2025

