

RECEIVED

By Tracy Crews at 10:59 am, Oct 16, 2023

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 119729	NAME OF AGENCY Newton County Sheriff's Office	DATE OF INSPECTION 10/16/2023
LOCATION OF INSTRUMENT (STREET AND CITY) 208 W Coler Street Neosho, MO 648502		TIME OF INSPECTION 10:10 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 23180 EXP. DATE 05/17/2025 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN DR 6930 SIM. NIST EXP DATE 07/24/2024 **CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 0.097

TEST 2 0.097

TEST 3 0.096

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Taylor Lombard

TYPE II PERMIT NUMBER/EXPIRATION DATE

230169 08/08/2025

TELEPHONE NUMBER

(417) 451-8300

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119729
Version no: 532C

TEST RECORD 00018

Temp Date Time 21^{9/}OL

Air Blank:
10/16/23 10:00 .000
Calibration Check:
21 10/16/23 10:00 .097

Subject Name

Test
Subject I.D.

Test 1
Operator Name, I.D.

Taylor Lombard 230169
Location
208 W Coler St
Neosho, MO 64850

AS IV Serial no: 119729
Version no: 532C

TEST RECORD 00019

Temp Date Time 21^{9/}OL

Air Blank:
10/16/23 10:03 .000
Calibration Check:
22 10/16/23 10:03 .097

Subject Name

Test
Subject I.D.

Test 2
Operator Name, I.D.

Taylor Lombard 230169
Location
208 W Coler St
Neosho, MO 64850

AS IV Serial no: 119729
Version no: 532C

TEST RECORD 00020

Temp Date Time 21^{9/}OL

Air Blank:
10/16/23 10:05 .000
Calibration Check:
23 10/16/23 10:05 .096

Subject Name

Test
Subject I.D.

Test 3
Operator Name, I.D.

Taylor Lombard 230169
Location
208 W Coler St
Neosho, MO
64850

AS IV Serial no: 119729
Version no: 532C

TEST RECORD 00021

Temp Date Time 21^{9/}OL

VOID: RFI
12 10/16/23 10:07

Subject Name

Test RFI
Subject I.D.

RFI
Operator Name, I.D.

Taylor Lombard 230169
Location
208 W Coler St
Neosho, MO 64850



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-864-8470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TAYLOR LOMBARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/8/2023

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230169

Paula J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/8/2025

MO 680-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LOMBARD, TAYLOR
Permit No 230169
Date Issued 8/8/2023 **Date Expires** 8/8/2025

