

**RECEIVED**

By Tracy Crews at 8:23 am, May 22, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005829	LOCATION OF INSTRUMENT JASPER CO. SHERIFF	DATE OF INSPECTION 04/02/2023	TIME OF INSPECTION 04:28
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	04:31	DRY	AG128602	10/13/2023
Cal Check	0.080	04:31	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	04:32	N/A	N/A	N/A
Cal Check	0.079	04:32	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	04:32	0.080	INTOXIMETERS	
Cal Check	0.080	04:33	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	04:33	0.080		
Cal Check	0.080	04:33	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	04:33	0.079		
<b>Pass</b>			CALIBRATION CHECK RESULT 3		
			0.080		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Result	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	04:34
RAM Test	Pass		Subject Test	RFI*	04:34
EEPROM Checksum Test	Pass		Air Blank	0.000	04:34
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
<b>Pass</b>			<b>Pass</b>		

**NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT**

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
1	0	1	7	4	1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <i>Scott M. A. #349</i>	PRINT NAME CHRISTOPHER CALVIN
TYPE II PERMIT NUMBER 220052	EXPIRATION DATE 02/19/2024
TELEPHONE NUMBER 4173588177	



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 13-Oct-2021

**Lot #** AG128602 **Model** 108

<b>Exp Date</b> 13-Oct-2023	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.080 ± 0.002 BrAC (208 ppm)
--------------------------------	-------------------------	---	--

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 10.13.2021 19:10

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**CHRISTOPHER CALVIN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/19/2022

NUMBER 220052

EXPIRES 2/19/2024

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CALVIN, CHRISTOPHER  
Permit No 220052  
Date Issued 2/19/2022    Date Expires 2/19/2024



JASPER COUNTY SHERIFF'S OFFICE

BREATH ALCOHOL PROGRAM -- INTOX 8000

FOR THE PERIOD BEGINNING 03-01-23 ENDING 04-02-23

MAINTENANCE DATE 03-01-23

DATE	SUSPECTS NAME	DOB	RESULT	OPERATOR
1 03/01/23	CLEVENGER, Phylcia	05/08/96	.106	MSHP TPR B.A. WALKER #412
2 3/6/23	Hannigan, Steven	9/22/72	.245	MSHP TPR T.C. Smith #1352
3 3/11/23	Howerton, Aaron	03/22/1987	.129	JCSO #351
4 3/12/23	Frazier, Kodie	8/11/93	.174	JCSO #361
5 3/15/23	Proficiency Tests X 38		.141	Calvin #349
6 3-18-23	Kennedy, Brittany	8-16-84	Def*	JCSO #335
7 3-26-23	Kidwell, Kyren	2-7-89	.085	CVPD #514
8 03/30/23	COBLE, ERMA V.	11/21/61	RFS	MSHP TPR B.A. WALKER #412
9 03/30/23	COBLE ERMA V.	11/21/61	<del>0.196</del>	MSHP TPR B.A. WALKER #412
10 04/01/23	Barnett, Leslie	8-29-74	.188	JCSO #335
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				
32				
33				
34				
35				
36				
37				
38				
39				
40				