

**RECEIVED**

By Brianna Medrano at 9:08 am, Apr 28, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005847	LOCATION OF INSTRUMENT IPD BATVAN STATION 1	DATE OF INSPECTION 04/24/2023	TIME OF INSPECTION 13:18
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG208102	STANDARD EXPIRATION DATE 03/22/2024
Air Blank	0.000	13:21	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.097	13:21	STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS, INC	
Air Blank	0.000	13:22	CALIBRATION CHECK RESULT 1 0.097		
Cal Check	0.095	13:22	CALIBRATION CHECK RESULT 2 0.095		
Air Blank	0.000	13:23	CALIBRATION CHECK RESULT 3 0.095		
Cal Check	0.095	13:24	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 5.0%	SPREAD (MUST BE .005 OR LESS) 0.002	
Air Blank	0.000	13:25	<b>Pass</b>		

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS		
Voltage/Current Test	Pass	Test	g/210L	Time
RAM Test	Pass	-----		
EEPROM Checksum Test	Pass	Air Blank	RFI*	13:25
Real Time Clock Test	Pass	Air Blank	0.000	13:26
DSP Test	Pass	*RFI Detect		
Analytical Stability Test	Pass	<b>Pass</b>		
Modem Test	Pass			
Temperature Regulation Test	Pass			

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>		
SIGNATURE 	PRINT NAME RYAN SCHILDKNECHT	
TYPE II PERMIT NUMBER 210253	EXPIRATION DATE 11/12/2023	TELEPHONE NUMBER 6605434573





STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**RYAN SCHILDKNECHT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2020

NUMBER 200133

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 580-0771 (6-10)

 **STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator SCHILDKNECHT, RYAN  
Permit No 200133  
Date Issued 3/9/2020 Date Expires 3/9/2022





**SERVICE WORK ORDER**

DATE: 12/20/2022

316 E 9th Street / Owensboro KY 42303 / USA  
Phone: 866-835-0690 Fax: 270-685-6268

**407449**

**Bill To:**  
Missouri Safety Center  
Central MO State University  
Humphreys Bldg Ste 200  
ATTN: ACCOUNTS PAYABLE  
Warrensburg, MO 64093  
USA  
Customer #: 640934  
Phone: 660-543-4830  
Fax: 660-543-4482

**Ship To:**  
Missouri Safety Center  
415 E Clark  
Marshall Bldg  
Warrensburg MO 64093  
USA  
Phone: 660-543-4830

**MODEL #:** 002480MO2

**Serial Number:** 80-005847

**BILL CODE:** Out of Warranty

**EXTRA PARTS RCVD:**

BH, KB, GAS BOX

**DESCRIPTION OF PROBLEM**

INTERNAL BATTERY AND PRINTING ISSUES

**WORK PERFORMED:**

Replaced parts listed; Calibrated/Final Tested; Software 8151.23.

**PARTS USED**

Seq. No.	Part	Description	Quantity
100	650517	CERTIFICATE OF CALIBRATION	2.00 EA
110	441169	COVER DUST,5/8in x 1/2in	1.00 EA
120	470145	CAP,PLASTIC,,.25IDx.50LG,RED	3.00 EA
130	470154	CAP,PLSTC,,.406IDx.438-.562,RED	1.00 EA
140	470218	SEAL,TAMPER EVIDENT,SERVICE	1.00 EA
150	470202	FILTER,FOAM,1.85x2.20	1.00 EA
160	530030	TUBING,TYGON,,.375IDx.562OD	2.00 IN
170	402705	SCR,M3.5x10 PAN HD,PHIL,SST	2.00 EA
180	408211	WSHR,LOCK,SPRING,SST,M3.5	2.00 EA
190	408206	WSHR,FLAT,SST,M4, 0.8mm THK	2.00 EA
200	402732	SCR,M3.5x12 PAN HD,PHIL,SST	1.00 EA
210	402725	SCR,M3x6 PAN HD,PHIL,SST	1.00 EA
220	690040	BATTERY,3VOLT,LITHIUM,CR2032	1.00 EA
230	210156	SENSOR, PRESSURE, 1000PSI	1.00 EA

**LABOR / TESTING**

Misc Code	Description	Hours
LABR	Service Repair Labor	
LCAL	Service - Calibration Adjust	0.50
LFT	Service - Cal / Final Test	1.00

Technician Name: Michael Byrne

SRV4074490001

