

RECEIVED

By Tracy Crews at 8:41 am, Sep 05, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005854	LOCATION OF INSTRUMENT LIBERTY POLICE DEPT.	DATE OF INSPECTION 09/03/2023	TIME OF INSPECTION 15:58
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG130104	STANDARD EXPIRATION DATE 10/28/2023
-----	-----	-----	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Air Blank	0.000	16:02	STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS	
Cal Check	0.100	16:02	CALIBRATION CHECK RESULT 1 0.100		
Air Blank	0.000	16:02	CALIBRATION CHECK RESULT 2 0.100		
Cal Check	0.100	16:03	CALIBRATION CHECK RESULT 3 0.100		
Air Blank	0.000	16:03	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 0.0%		
Cal Check	0.100	16:04	SPREAD (MUST BE .005 OR LESS) 0.000		
Air Blank	0.000	16:04			
Pass					

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----	-----	-----
EEPROM Checksum Test	Pass		Air Blank	0.000	16:05
Real Time Clock Test	Pass		Subject Test	RFI*	16:05
DSP Test	Pass		Air Blank	0.000	16:05
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass		Pass		
Temperature Regulation Test	Pass				
Pass					

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
1	0	0	0	0	0	1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

NONE

INSPECTING OFFICER			
SIGNATURE 	PRINT NAME BENJAMIN LAUGHLIN	TYPE II PERMIT NUMBER 230106	EXPIRATION DATE 05/30/2025
		TELEPHONE NUMBER 8164394701	