



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

RECEIVED  
BY 7/18/23 10:41 AM

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005855	LOCATION OF INSTRUMENT CARTHAGE POLICE	DATE OF INSPECTION 09/01/2023	TIME OF INSPECTION 02:27
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	02:28	DRY	AG222301	08/11/2024
Cal Check	0.099	02:29	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	02:29	N/A	N/A	N/A
Cal Check	0.099	02:30	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	02:30	0.100	INTOXIMETERS	
Cal Check	0.100	02:30	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	02:31	0.099		
<h1>Pass</h1>			CALIBRATION CHECK RESULT 2		
			0.099		
			CALIBRATION CHECK RESULT 3		
			0.100		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.0%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		-----		
RAM Test	Pass		Air Blank	0.000	02:32
EEPROM Checksum Test	Pass		Subject Test	RFI*	02:32
Real Time Clock Test	Pass		Air Blank	0.000	02:33
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass		<h1>Pass</h1>		
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT					
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

AUGUST 2023 MAINTENANCE

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME BRADY LUCE	
TYPE II PERMIT NUMBER 220135	EXPIRATION DATE 05/11/2024	TELEPHONE NUMBER 4172377200



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
Exclusive Supplier  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 16-Aug-2022

**Lot #** AG222301 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
11-Aug-2024	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by: Quality Control  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)  
Date: 08.18.2022 21:07

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRADY LUCE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220135

EXPIRES 5/11/2024

*Laura G. Day*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator LUCE, BRADY  
 Permit No 220135  
 Date Issued 5/11/2022 Date Expires 5/11/2024

