



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER <b>80-006195</b>	LOCATION OF INSTRUMENT <b>CARTHAGE POLICE DEPT</b>	DATE OF INSPECTION <b>02/04/2023</b>	TIME OF INSPECTION <b>03:20</b>
--	---	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE <b>DRY</b>	STANDARD LOT # <b>AG222301</b>	STANDARD EXPIRATION DATE <b>08/11/2024</b>
-----	-----	-----	SIM TEMPERATURE <b>N/A</b>	SIM SERIAL NUMBER <b>N/A</b>	SIM CERTIFICATE EXPIRATION <b>N/A</b>
Air Blank	0.000	03:24	STANDARD VALUE <b>0.100</b>	STANDARD SUPPLIER <b>INTOXIMETERS</b>	
Cal Check	0.099	03:24	CALIBRATION CHECK RESULT 1 <b>0.099</b>		
Air Blank	0.000	03:25	CALIBRATION CHECK RESULT 2 <b>0.103</b>		
Cal Check	0.103	03:25	CALIBRATION CHECK RESULT 3 <b>0.102</b>		
Air Blank	0.000	03:26	MAXIMUM DEVIATION (MUST BE WITHIN 5%) <b>3.0%</b>		
Cal Check	0.102	03:26	SPREAD (MUST BE .005 OR LESS) <b>0.004</b>		
Air Blank	0.000	03:26	<h1>Pass</h1>		
Air Blank	0.000	03:26			

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----	-----	-----
EEPROM Checksum Test	Pass		Air Blank	0.000	03:27
Real Time Clock Test	Pass		Subject Test	RFI*	03:27
DSP Test	Pass		Air Blank	0.000	03:28
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass		<h1>Pass</h1>		
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
**JAN 2023 MAINT.**

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME <b>RANSOME</b>	
TYPE II PERMIT NUMBER <b>220138</b>		EXPIRATION DATE <b>05/11/2024</b>	TELEPHONE NUMBER <b>4172377200</b>



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 16-Aug-2022

**Lot # AG222301 Model 108**

<b>Exp Date</b> 11-Aug-2024	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (260 ppm)
--------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:08.18.2022 21:07

**Approved for Release:**                     *Rod Marsala*                      
 Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JACCOB RANSOME**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220138

EXPIRES 5/11/2024

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** RANSOME, JACCOB  
**Permit No** 220138  
**Date Issued** 5/11/2022    **Date Expires** 5/11/2024