



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-006195	LOCATION OF INSTRUMENT CARTHAGE POLICE DEPT	DATE OF INSPECTION 03/02/2023	TIME OF INSPECTION 04:59
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG222301	STANDARD EXPIRATION DATE 08/11/2024
-----	-----	-----	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Air Blank	0.000	05:00	STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS	
Cal Check	0.102	05:01	CALIBRATION CHECK RESULT 1 0.102		
Air Blank	0.000	05:01	CALIBRATION CHECK RESULT 2 0.102		
Cal Check	0.102	05:02	CALIBRATION CHECK RESULT 3 0.102		
Air Blank	0.000	05:02	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.0%		
Cal Check	0.102	05:02	SPREAD (MUST BE .005 OR LESS) 0.000		
Air Blank	0.000	05:03			

Pass

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS	
Voltage/Current Test	Pass	Test	g/210L
RAM Test	Pass	-----	-----
EEPROM Checksum Test	Pass	Air Blank	0.000
Real Time Clock Test	Pass	Subject Test	RFI*
DSP Test	Pass	Air Blank	0.000
Analytical Stability Test	Pass		
Modem Test	Pass	*RFI Detect	
Temperature Regulation Test	Pass		

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
FEBRUARY 2023 MAINTENANCE

INSPECTING OFFICER		
SIGNATURE <i>Brady Luce</i>	PRINT NAME BRADY LUCE	
TYPE II PERMIT NUMBER 220135	EXPIRATION DATE 05/11/2024	TELEPHONE NUMBER 4172377200



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRADY LUCE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

Laura A. Natz

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220135

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/11/2024

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LUCE, BRADY
 Permit No 220135
 Date Issued 5/11/2022 Date Expires 5/11/2024

