

RECEIVED

By Tracy Crews at 4:00 pm, Feb 17, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

SC PD

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007511	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 02/07/2023	TIME OF INSPECTION 01:04
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	01:05	DRY	33321080A3	01/05/2024
Cal Check	0.079	01:05	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	01:06	N/A	N/A	N/A
Cal Check	0.080	01:06	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	01:07	0.080	CMI INC	
Cal Check	0.080	01:07	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	01:08	0.079		
Cal Check	0.080	01:08	CALIBRATION CHECK RESULT 2		
Pass			0.080		
			CALIBRATION CHECK RESULT 3		
			0.080		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	01:08
RAM Test	Pass		Air Blank	0.000	01:09
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass		Pass		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
3	0	2	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT WAS TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE	PRINT NAME	TYPE II PERMIT NUMBER	EXPIRATION DATE
	JORDAN INFRANCA	210128	06/21/2023
		TELEPHONE NUMBER	816-382-5897



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-245-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 14496
Part #: BAC135L080T
Cylinder Size: 105L
Lot Number: 33321080A3
Expiration: 1/5/2024

0.080 BAC for the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration:	Analytical Accuracy (U, ke2):	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/- 0.882 acc(±28.1)	NDIR	CM1 Inc
Nitrogen	Balance	[± 2 ppm]		316 East Ninth Street Cynthiana, KY 42303 Phone 856-835-0690 www.alcoholtest.com

*Traceable to:
Certified Reference Material - 2624 (µmol/mol Ethanol in Nitrogen - Serial No. G10015026 Lot No. 050319E11

Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

DL-04-2022
Instance Data



The operation remains within the certificate was obtained using equipment and methods capable of producing analytical results traceable to NIST and apply only to the items checked on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information has a certain level of accuracy and risk of this data. Liability shall be limited to established replacement cost of the material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):
ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/21/2021
NUMBER: 210128
EXPIRES: 6/21/2023
MO 888-0771 (6-18)
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAD 098/01
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The named certificate holder is authorized to operate an alcohol breath analyzer in Missouri for the determination of the alcoholic content of breath from a sample of expired air.

Operator: INFRANCA, JORDAN
Permit No: 210128
Date issued: 6/21/2021
Date Expires: 6/21/2023