

RECEIVED

By Tracy Crews at 7:27 am, Aug 25, 2023

Scpd



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007511	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 08/13/2023	TIME OF INSPECTION 22:12
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:14	DRY	33321080A3	01/05/2024
Cal Check	0.078	22:14	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:15	N/A	N/A	N/A
Cal Check	0.079	22:15	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:16	0.080	CMI INC	
Cal Check	0.079	22:16	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:16	0.078		
Cal Check	0.079	22:16	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	22:16	0.079		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	22:17
RAM Test	Pass		Air Blank	RFI*	22:17
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass				
DSP Test	Pass				
Analytical Stability Test	Pass				
Internal Printer Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	75	1	0	0	3	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

BREATH INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE 	PRINT NAME INFRANCA, JORDAN	
TYPE II PERMIT NUMBER 230118	EXPIRATION DATE 06/07/2025	TELEPHONE NUMBER 816-382-5897

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62661-0790
 217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com



Certificate of Analysis

Certificate ID: 14496
 Part #: BAC105L088T
 Cylinder Size: 105L
 Lot Number: 33321080A3
 Expiration: 1/5/2024

0.080 BA C (For the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration: (U, k=2):	Analytical Accuracy	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/-0.882 BAC(G/210)	NDIR	CMI Inc 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com
Nitrogen	Balance	(5.2 ppm)		

Traceable to:
 Certified Reference Material - 2024.4 µmol/mol
 Ethanol in Nitrogen - Serial No. GN0015026 Loc No. 050319E11

Score in dry area, away from sources of heat,
 ignition and direct sunlight. Do not allow storage
 area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

01-04-2022
 Issuance Date



The information on this certificate was obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the stated purpose. The information is not to be used for any other purpose, and the user is responsible for the use of any information provided for any particular purpose. The information is not to be used for any other purpose, and the user is responsible for the use of any information provided for any particular purpose.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

PERMIT
 TYPE II
JORDAN INFRANCA



is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/7/2023
 NUMBER: 230118
 EXPIRES: 6/7/2025
 MO 980-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
 LMS 08-610

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD
 The number contained in this instrument is authorized to operate an Intoxilizer breath alcohol analyzer for the determination of the alcoholic content of breath from a sample of expired air in Missouri.
 OPERATOR: INFRANCA, JORDAN
 Permit No: 230118
 Date Issued: 6/7/2023
 Date Expires: 6/7/2025