

RECEIVED

By Tracy Crews at 2:51 pm, Nov 08, 2023

SCPD



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007511	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 10/12/2023	TIME OF INSPECTION 11:22
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 33321080A3	STANDARD EXPIRATION DATE 01/05/2024
Air Blank	0.000	11:24	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.078	11:24	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI INC	
Air Blank	0.000	11:25	CALIBRATION CHECK RESULT 1 0.078		
Cal Check	0.079	11:25	CALIBRATION CHECK RESULT 2 0.079		
Air Blank	0.000	11:26	CALIBRATION CHECK RESULT 3 0.079		
Cal Check	0.079	11:26	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.5%	SPREAD (MUST BE .005 OR LESS) 0.001	
Air Blank	0.000	11:26	Pass		

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS	
Voltage/Current Test	Pass	Test	g/210L
RAM Test	Pass	Air Blank	RFI*
EEPROM Checksum Test	Pass	Air Blank	0.000
Real Time Clock Test	Pass	*RFI Detect	
DSP Test	Pass		
Analytical Stability Test	Pass		
Modem Test	Pass		
Temperature Regulation Test	Pass		
	Pass		Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	00-04	.05-09	.10-14	.15-19	OVER .19	
0	0	0	2	1	0	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

BREATH INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE <i>P.O. J. Infranca</i>	PRINT NAME INFRANCA, JORDAN
--------------------------------------	--------------------------------

TYPE II PERMIT NUMBER 230118	EXPIRATION DATE 06/07/2025	TELEPHONE NUMBER 816-382-5897
---------------------------------	-------------------------------	----------------------------------



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62451-0790
 217-245-2733 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 14496
 Part #: BAC1051080T
 Cylinder Size: 105L
 Lot Number: 3332-080A3
 Expiration: 1/5/2024

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, 1σ)	Analytical Method:	Distributed by:
Ethanol	248 ppm Balance	±0.003 3σ(671sk): NOTR		CMJ Inc 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

*Traceable to: Certified Reference Material - 2624 µmol/mol Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050315E11
 Scores in dry area, away from sources of heat, ignition and direct sunlight. DO NOT allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
 Issued Date: 01-04-2022
 Accreditation: 151189

The accuracy results shown on this certificate were obtained using techniques and methods capable of producing analytical results traceable to NIST, and apply only to the item indicated on this certificate. LIND Products Company makes no warranty or representation as to the suitability of this gas for any information provided for any particular purpose. The information on this certificate is for informational purposes only and is not intended to constitute a contract or any other legal instrument.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH-ALCOHOL PROGRAM

**PERMIT
 TYPE II**

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXLYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/7/2023
 NUMBER: 230118
 EXPIRES: 6/7/2025
 MO-590-071 (9-0)

Mike Morrison
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Doreen J. Nielsen
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH-ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The named contributor is authorized to operate an educational breath alcohol instrument for the determination of the alcoholic content in breath from an operator at Missouri.

INFRANCA, JORDAN
 Permit No. 230118
 Date Issued 6/7/2023 Date Expires 6/7/2025