

RECEIVED

By Tracy Crews at 9:18 am, Feb 14, 2023

m pd.



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007512		LOCATION OF INSTRUMENT KANSAS CITY POLICE		DATE OF INSPECTION 02/06/2023		TIME OF INSPECTION 22:23	
CALIBRATION CHECK RESULTS				CALIBRATION CHECK SUMMARY			
Test	g/210L	Time		STANDARD TYPE DRY	STANDARD LOT # 33321080A3	STANDARD EXPIRATION DATE 01/05/2024	
Air Blank	0.000	22:25		SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A	
Cal Check	0.079	22:25		STANDARD VALUE 0.080	STANDARD SUPPLIER CMI INC		
Air Blank	0.000	22:25		CALIBRATION CHECK RESULT 1 0.079			
Cal Check	0.079	22:26		CALIBRATION CHECK RESULT 2 0.079			
Air Blank	0.000	22:26		CALIBRATION CHECK RESULT 3 0.079			
Cal Check	0.079	22:27		MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		SPREAD (MUST BE .005 OR LESS) 0.000	
Air Blank	0.000	22:27					
Pass							
DIAGNOSTIC TEST RESULTS				RFI TEST RESULTS			
Voltage/Current Test	Pass	Pass		Test	g/210L	Time	
RAM Test	Pass	Pass		Air Blank	RFI*	22:27	
EEPROM Checksum Test	Pass	Pass		Air Blank	0.000	22:28	
Real Time Clock Test	Pass	Pass		*RFI Detect			
DSP Test	Pass	Pass					
Analytical Stability Test	Pass	Pass					
Modem Test	Pass	Pass					
Temperature Regulation Test	Pass	Pass					
Pass				Pass			
NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	2	1	2	2	9		
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary). INSTRUMENT WAS TESTED AND CERTIFIED							
INSPECTING OFFICER							
SIGNATURE <i>P.O. J. Infranca</i>				PRINT NAME JORDAN INFRANCA			
TYPE II PERMIT NUMBER 210128		EXPIRATION DATE 06/21/2023			TELEPHONE NUMBER 816-382-5897		



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-243-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 14496
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 33321080A3
Expiration: 1/5/2024

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration	Analytical Accuracy (U, L=2)	Analytical Method
Ethanol	208 ppm	±0.002 BAC (0.210L)	NDIR
Nitrogen	Balance	±0.2 ppm	

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 856-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319B11

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

01-04-2022
Issuance Date



The calibration return upon this certificate was obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT TYPE II JORDAN INFRANCA

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcohol content of blood from a sample of expired air. Permit issued under the provisions of sections 877.020 through 877.041, RSMo and 806.111 through 806.119 RSMo.

DATE 6/21/2021

Laura P. Wang

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210128

EXPIRES 6/21/2023

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 880-0771 (8-10)

LAB-4 (06-10)

STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES BREATH ALCOHOL PROGRAM	
INSTRUMENT OPERATOR CARD	
<small>The named operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcohol content in breath from an expired air in Missouri.</small>	
Operator	INFRANCA, JORDAN
Permit No	210128
Date Issued	6/21/2021 Date Expires 6/21/2023