

RECEIVED

By Tracy Crews at 2:51 pm, Nov 08, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007512	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 10/06/2023	TIME OF INSPECTION 13:11
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	13:13	DRY	33321080A3	01/05/2024
Cal Check	0.079	13:13	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	13:14	N/A	N/A	N/A
Cal Check	0.080	13:14	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	13:15	0.080	CMI INC	
Cal Check	0.079	13:15	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	13:15	0.079		
Cal Check	0.079	13:15	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	13:15	0.080		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Result	Time	Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	13:16
RAM Test	Pass		Subject Test	RFI*	13:17
EEPROM Checksum Test	Pass		Air Blank	0.000	13:17
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass		Pass		
Modem Test	Pass		Pass		
Temperature Regulation Test	Pass		Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT					
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
Time-Date changed.

INSPECTING OFFICER		SIGNATURE		PRINT NAME	
				WADE ROBINSON	
TYPE II PERMIT NUMBER		EXPIRATION DATE		TELEPHONE NUMBER	
210266		11/18/2023		8164828141	



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



WADE ROBINSON

**PERMIT
TYPE II**

7 Eastgate Dr. • P.O. Box 790 • Jacksonvillle, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com



Certificate of Analysis

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):
ALCO-SENSOR IV WITH PRINTER, INTOXLYZER 8000
for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/18/2021
NUMBER 210266
EXPIRES 11/18/2023
MOS 580.071 (6-10)

Deanna Q. Day
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
David A. Robinson
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LMS (89-40)



Certificate ID: 14496
Part #: BAC1051080T
Cylinder Size: 105L
Lot Number: 33321080A3
Expiration: 1/5/2024

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, L=2):	Analytical Method:
Ethanol	268 ppm	+/- 0.8% BAC (C/2AB)	NDIR
Nitrogen	Balance	(3.2 ppm)	

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 2624 µmol/mol
Ethanol in Nitrogen - Serial No. GNI0015026 Lot No. 050319E11

Score in d/r area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Wade Robinson
Specialty Gas Lab Tech

01-04-2022
Issuance Date



The information on this certificate was obtained using equipment and methods capable of producing analytical results traceable to NIST and apply only to the sample. The information on the label does not constitute a warranty. Liability shall be limited to replacement costs of this instrument or service.

ISO/IEC 17025:2017 Accredited Laboratory