

RECEIVED

By Tracy Crews at 4:00 pm, Feb 17, 2023

mpd.



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007512	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 02/06/2023	TIME OF INSPECTION 22:23
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:25	DRY	33321080A3	01/05/2024
Cal Check	0.079	22:25	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:25	N/A	N/A	N/A
Cal Check	0.079	22:26	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:26	0.080	CMI INC	
Cal Check	0.079	22:27	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:27	0.079		
Cal Check	0.079	22:27	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	22:27	0.079		
<b>Pass</b>			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	RFI*	22:27
RAM Test	Pass		Air Blank	0.000	22:28
EEPROM Checksum Test	Pass		*RFI Detect		
Real Time Clock Test	Pass		<b>Pass</b>		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
<b>Pass</b>			<b>Pass</b>		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	2	1	2	2	9		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
INSTRUMENT WAS TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE <i>P.O. J. Infranca</i>		PRINT NAME JORDAN INFRANCA	
TYPE II PERMIT NUMBER 210128	EXPIRATION DATE 06/21/2023	TELEPHONE NUMBER 816-382-5897	



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2193 • Fax: 217-243-7634 • www.ilmoproducts.com

### Certificate of Analysis

Certificate ID: 14496  
Part #: BAC105L080T  
Cylinder Size: 105L  
Lot Number: 33321088A3  
Expiration: 1/5/2024

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration	Analytical Accuracy (U, kcal%)	Analytical Method	Distributed by:
Ethanol	288 ppm	+/- 0.002 BAC (± 2%)	NDIR	CMI, Inc. 316 East Ninth Street Covington, KY 42303 Phone 866-835-0690 www.cmihoholtest.com
Nitrogen	Balance			

\*Traceable to Certified Reference Material - 2624 (umol/mol Ethanol in Nitrogen - Serial No. G10015026 Lot No. 050319E11)

Score in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech  
*[Signature]*

DL-04-2022  
Insurance Date



ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
TYPE II  
JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):  
**ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 6/21/2021  
NUMBER: 210128  
EXPIRES: 6/21/2023  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
LAW 198-01

*[Signature]*  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM  
INSTRUMENT OPERATOR CARD

The primary responsibility of an operator is to ensure an individual having alcohol in their system is not permitted to operate a motor vehicle.  
Operator: INFRANCA, JORDAN  
Permit No: 210128  
Date issued: 6/21/2021  
Date Expires: 6/21/2023

The expiration date from this certificate was obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items covered on this certificate. LMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information on this site is provided as a service to our customers and is not intended to constitute an offer of any product or service.