

RECEIVED

By Tracy Crews at 1:39 pm, Jul 21, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007513	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 07/14/2023	TIME OF INSPECTION 08:52
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	08:54	DRY	33321080A3	01/05/2024
Cal Check	0.078	08:54	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	08:54	N/A	N/A	N/A
Cal Check	0.079	08:55	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	08:55	0.080	CMI INC	
Cal Check	0.078	08:56	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	08:56	0.078		
Cal Check	0.078	08:56	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	08:56	0.079		
Pass			CALIBRATION CHECK RESULT 3		
			0.078		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	08:58
RAM Test	Pass		Subject Test	RFI*	08:59
EEPROM Checksum Test	Pass		Air Blank	0.000	08:59
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass				
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	1	0	3	1	0	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
Time-Date changed.

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME WADE ROBINSON	
TYPE II PERMIT NUMBER 210266	EXPIRATION DATE 11/18/2023	TELEPHONE NUMBER 8164828141



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

WADE ROBINSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/18/2021
NUMBER 210266
EXPIRES 11/18/2023

Laura A. Wray
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kaunz
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 98-0771 (5-19)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air.

Operator: **ROBINSON, WADE**
Permit No: **210266**
Date Issued: **11/18/2021** Date Expires: **11/18/2023**



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 14496
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 33321080A3
Expiration: 1/5/2024

0.080 BAC (for the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Accuracy (U, L):	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/-0.002 BAC(0.210L)	NDIR	CMI Inc.
Nitrogen	Balance	(5.3 ppm)		316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Johnnie
Spec. Gas Lab Tech

J. Clift

01-04-2022
Issuance Date



The analysis results of this certificate were obtained using equipment and standards capable of producing laboratory results traceable to NIST. We apply only to the items identified on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of use of any information provided for any particular purpose. The information is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of the material or service.

ISO/IEC 17025:2017 Accredited Laboratory