

RECEIVED

By Tracy Crews at 7:27 am, Aug 25, 2023

Spd.



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007513	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 08/13/2023	TIME OF INSPECTION 18:05
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	18:07
Cal Check	0.079	18:07
Air Blank	0.000	18:08
Cal Check	0.079	18:08
Air Blank	0.000	18:09
Cal Check	0.080	18:09
Air Blank	0.000	18:10

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE DRY	STANDARD LOT # 33321080A3	STANDARD EXPIRATION DATE 01/05/2024
SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
STANDARD VALUE 0.080	STANDARD SUPPLIER CMI INC	
CALIBRATION CHECK RESULT 1 0.079		
CALIBRATION CHECK RESULT 2 0.079		
CALIBRATION CHECK RESULT 3 0.080		
MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		SPREAD (MUST BE .005 OR LESS) 0.001

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	18:10
Reference	RFI*	18:10
Air Blank	0.000	18:11

*RFI Detect

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	43	0	0	0	2

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

BREATH INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE <i>P.O. J. Infranca</i>	PRINT NAME INFRANCA, JORDAN
--------------------------------------	--------------------------------

TYPE II PERMIT NUMBER 230118	EXPIRATION DATE 06/07/2025	TELEPHONE NUMBER 816-382-5897
---------------------------------	-------------------------------	----------------------------------



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 14496

Part #: BAC1051080T

Cylinder Size: 105L

Lot Number: 33321080A3

Expiration: 1/5/2024

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Component	Reported Concentration	Analytical Accuracy (U, k=2)	Analytical Method
Ethanol	288 ppm	+/- 0.002 BAC(G/200L)	NDIR
Nitrogen	Balance	[3.2 ppm]	

Distributed by:
CMI Inc
316 East Ninth Street
Owensboro, KY 42303
Phone 866-833-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 umol/lmol
Ethanol in Nitrogen - Serial No. GNM0015025 Lot No. 0501319E11

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

01-04-2022
Issuance Date



The opinion results from this certificate were obtained using equipment and reagents capable of producing analytical results traceable to NIST, and apply only to the items
person. The information used to determine the accuracy of the results is the responsibility of the user. The information used to determine the accuracy of the results is the responsibility of the user.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

JORDAN INFRANCA



is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

M. De M...
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Jordan Infranca
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

DATE: 6/7/2023
NUMBER: 230118
EXPIRES: 6/7/2025
MO 960-0771 (6-10)

STATE OF MISSOURI SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The instrument operator is responsible to operate an instrument during sections
in Missouri.

Operator: INFRANCA, JORDAN
Permit No: 230118
Date Issued: 6/7/2023 Date Expires: 6/7/2025

UM-4 (9-10)