



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMi INTOXILYZER 8000 MAINTENANCE REPORT

EPD

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

| | | | |
|---------------------------------------|----------------------------------------------|----------------------------------|-----------------------------|
| INSTRUMENT SERIAL NUMBER 80-007514 | LOCATION OF INSTRUMENT KANSAS CITY POLICE | DATE OF INSPECTION 02/06/2023 | TIME OF INSPECTION 23:33 |
|---------------------------------------|----------------------------------------------|----------------------------------|-----------------------------|

| CALIBRATION CHECK RESULTS | | | CALIBRATION CHECK SUMMARY | | |
|---------------------------|--------|-------|-----------------------------------------------|------------------------------|----------------------------------------|
| Test | g/210L | Time | STANDARD TYPE DRY | STANDARD LOT # 33321080A3 | STANDARD EXPIRATION DATE 01/05/2024 |
| Air Blank | 0.000 | 23:35 | SIM TEMPERATURE N/A | SIM SERIAL NUMBER N/A | SIM CERTIFICATE EXPIRATION N/A |
| Cal Check | 0.077 | 23:35 | STANDARD VALUE 0.080 | STANDARD SUPPLIER CMI INC | |
| Air Blank | 0.000 | 23:36 | CALIBRATION CHECK RESULT 1 0.077 | | |
| Cal Check | 0.076 | 23:36 | CALIBRATION CHECK RESULT 2 0.076 | | |
| Air Blank | 0.000 | 23:36 | CALIBRATION CHECK RESULT 3 0.077 | | |
| Cal Check | 0.077 | 23:37 | MAXIMUM DEVIATION (MUST BE WITHIN 5%) 5.0% | | |
| Air Blank | 0.000 | 23:37 | SPREAD (MUST BE .005 OR LESS) 0.001 | | |

Pass

| DIAGNOSTIC TEST RESULTS | | | RFI TEST RESULTS | | |
|-----------------------------|------|--|------------------|--------|-------|
| Voltage/Current Test | Pass | | Test | g/210L | Time |
| RAM Test | Pass | | Air Blank | RFI* | 23:37 |
| EEPROM Checksum Test | Pass | | Air Blank | 0.000 | 23:38 |
| Real Time Clock Test | Pass | | *RFI Detect | | |
| DSP Test | Pass | | Pass | | |
| Analytical Stability Test | Pass | | Pass | | |
| Modem Test | Pass | | Pass | | |
| Temperature Regulation Test | Pass | | Pass | | |

Pass

Pass

| NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT | | | | | | |
|-----------------------------------------------------------------------------------------|---------|---------|---------|---------|----------|--|
| REFUSALS | .00-.04 | .05-.09 | .10-.14 | .15-.19 | OVER .19 | |
| 1 | 3 | 0 | 0 | 0 | 6 | |

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT TESTED AND CERTIFIED

| | | |
|---------------------------------|-------------------------------|----------------------------------|
| INSPECTING OFFICER | | |
| SIGNATURE | PRINT NAME JORDAN INFRANCA | |
| TYPE II PERMIT NUMBER 210128 | EXPIRATION DATE 06/21/2023 | TELEPHONE NUMBER 816-382-5897 |



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2153 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 14496
Part #: BAC185L080T
Cylinder Size: 185L
Lot Number: 33321080A3
Expiration: 1/5/2024

0.080 SAC (For the calibration of instruments used to determine breath alcohol concentration)

| Components | Reported Concentration | Accuracy (U, H2): | Analytical Method |
|------------|------------------------|--------------------------|-------------------|
| Ethanol | 288 ppm | +/-0.003, SAC(72A), ND7A | |
| Nitrogen | Balance | [5, 2 ppm] | |

Distributed by:
CMI, Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 856-835-0690
www.alcoholtest.com

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Score in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech
[Signature]

01-04-2022
Instance Data



ISO/IEC 17025:2017 Accredited Laboratory

The operation of this system was obtained using equipment and materials capable of producing analytical results suitable to NIST, and apply only to the item
analyzed at this facility. LMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular
purpose. This information is to be used in accordance with the instructions and safety data sheets for the equipment and materials used in the analysis of the item.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT TYPE II JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
and operate the following breath analyzer(s):
ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/21/2021
NUMBER 210128
EXPIRES 6/21/2023
Mo 588-9771 (8-15)
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
[Signature]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LMS-PR-01

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The holder of this card is authorized to operate an Intoxilizer 8000 Alcolyzer in Missouri.
Operator: INFRANCA, JORDAN
Permit No: 210128
Date Issued: 06/21/2021 Date Expires: 06/21/2023