

RECEIVED

By Tracy Crews at 7:27 am, Aug 25, 2023



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007514	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 08/13/2023	TIME OF INSPECTION 21:35
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	21:36	DRY	33321080A3	01/05/2024
Cal Check	0.079	21:37	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	21:37	N/A	N/A	N/A
Cal Check	0.079	21:38	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	21:38	0.080	CMI INC	
Cal Check	0.078	21:38	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	21:39	0.079		
Pass			CALIBRATION CHECK RESULT 2		
			0.079		
			CALIBRATION CHECK RESULT 3		
			0.078		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			2.5%	0.001	

DIAGNOSTIC TEST RESULTS		RFI TEST RESULTS	
		Test	g/210L
Voltage/Current Test	Pass		Time
RAM Test	Pass		
EEPROM Checksum Test	Pass	Air Blank	RFI*
Real Time Clock Test	Pass	Air Blank	0.000
DSP Test	Pass		21:39
Analytical Stability Test	Pass		21:40
Modem Test	Pass	*RFI Detect	
Temperature Regulation Test	Pass		
Pass		Pass	

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	0	4	2	5	1		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
BREATH INSTRUMENT TESTED AND CERTIFIED

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME INFRANCA, JORDAN
TYPE II PERMIT NUMBER 230118	EXPIRATION DATE 06/07/2025
TELEPHONE NUMBER 816-382-5897	

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
 217-245-2183 • Fax: 217-243-7634 • WWW.ILMOPRODUCTS.COM



Certificate of Analysis

Certificate ID: 14496
 Part #: BAC105L080T
 Cylinder Size: 105L
 Lot Number: 33321080A3
 Expiration: 1/5/2024

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)
 Analytical
 Reported Accuracy (U, kcal): Analytical
 Concentration: Method:
 Component: 200 ppm Balance
 Ethanol: ±0.002 BAC(62air) NGIR
 Nitrogen: ±0.2 ppm

Distributed by:
 CMI Inc
 316 East Ninth Street
 Owensboro, KY 42303
 Phone 866-835-0690
 www.alcoholtest.com

Traceable to:
 Certified Reference Material - 2624 µmol/lmol
 Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Score in dry area, away from sources of heat,
 ignition and direct sunlight. Do not allow storage
 area to exceed 52 °C (125 °F).

01-04-2022
 Issuance Date

Specialty Gas Lab Tech



The operator reads from this certificate were obtained using equipment and procedures capable of producing analytical results traceable to NIST, and apply only to the items
 described herein. The information is not intended to be used for any other purpose. The information is not intended to be used for any other purpose. The information is not intended to be used for any other purpose.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
 and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
 577.020 through 577.041, RSMo and 306.11 through 306.119 RSMo.

DATE: 6/7/2023
 NUMBER: 230118
 EXPIRES: 6/7/2025
 MAC: 986-0771 (6-10)
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
 Director of Department of Health and Senior Services
 L&L (98-10)

STATE OF MISSOURI SENIOR SERVICES
 BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD
 This permit card is intended to identify an instrument operator in Missouri.
 Operator: INFRANCA, JORDAN
 Permit No: 230118
 Date Issued: 6/7/2023 Data Expires: 6/7/2025