



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

REPORT #7

By Tracy Crews at 7:34 am, Jan 29, 2024

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>030807</b>	NAME OF AGENCY <b>Jefferson County Sheriff's Office</b>	DATE OF INSPECTION <b>01/25/2024</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>26 Dillon Plaza High Ridge, MO 63049 (JCSO N. Zone Office)</b>		TIME OF INSPECTION <b>1822</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeter LOT # AG305902 EXP. DATE 02/28/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← **.098**

TEST 2 ← **.097**

TEST 3 ← **.096**

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE *Zachary Reed*

PRINT NAME **Zachary Reed #810**

TYPE II PERMIT NUMBER/EXPIRATION DATE **230236 / 10/31/2025**

TELEPHONE NUMBER  
(636 ) 797-5000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 030807  
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00356

Temp Date Time 21<sup>9/</sup>OL

Air Blank:  
01/25/24 18:22 .000  
Calibration Check:  
-18 01/25/24 18:22 .098

Subject Name

Test 1  
Subject I.D.

Operator Name, I.D.

Dr Z Reed 810/230236  
Location

AS IV Serial no: 030807  
Version no: 532B

TEST RECORD 00357

Temp Date Time 21<sup>9/</sup>OL

Air Blank:  
01/25/24 18:28 .000  
Calibration Check:  
20 01/25/24 18:28 .097

Subject Name

Test 2  
Subject I.D.

Operator Name, I.D.

Dr Z Reed 810/230236  
Location

AS IV Serial no: 030807  
Version no: 532B

TEST RECORD 00358

Temp Date Time 21<sup>9/</sup>OL

Air Blank:  
01/25/24 18:30 .000  
Calibration Check:  
21 01/25/24 18:30 .096

Subject Name

Test 3  
Subject I.D.

Operator Name, I.D.

Dr Z Reed 810/230236  
Location

AS IV Serial no: 030807  
Version no: 532B

TEST RECORD 00359

Temp Date Time 21<sup>9/</sup>OL

VOID: RFI  
12 01/25/24 18:32

Subject Name

Test 4 RFI  
Subject I.D.

Operator Name, I.D.

Dr Z Reed 810/230236  
Location



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 28-Feb-2023

**Lot #** AG305902 **Model** 108

<b>Exp Date</b> 28-Feb-2025	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:03.01.2023 17:25

**Approved for Release:**   
 Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**