



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 035714	PRINTER SN 092.3576.249	DATE OF INSPECTION 03/26/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 203 Veterans Memorial Parkway, Wright City, Missouri, 63390		TIME OF INSPECTION 1:28 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 23390 EXP. DATE 10/17/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0C SIMULATOR SN SD2745 SIMULATOR EXP DATE 05/10/2024

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102	TEST 2 .102	TEST 3 .102
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
 Time updated due to day light savings.

INSPECTING OFFICER

SIGNATURE 104	PRINT NAME G. Monroe
TYPE II PERMIT NUMBER/EXPIRATION DATE 230040 03/09/2025	TELEPHONE NUMBER (636) 745-3541

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number **FN03072301** whose values are traceable to NIST.*

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 035714
Version no: 5328

TEST RECORD 01558

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:28 .000

Calibration Check: 24 03/26/24 13:28 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Monroe 104 230040

Operator Name: I.D.

Wrisant City PD

Location

AS IV Serial no: 035714
Version no: 5328

TEST RECORD 01559

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:34 .000

Calibration Check: 24 03/26/24 13:34 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Monroe 104 230040

Operator Name: I.D.

Wrisant City PD

Location

AS IV Serial no: 035714
Version no: 5328

TEST RECORD 01560

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:38 .000

Calibration Check: 25 03/26/24 13:38 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Monroe 104 230040

Operator Name: I.D.

Wrisant City PD

Location

AS IV Serial no: 035714
Version no: 5328

TEST RECORD 01561

Temp Date Time 210L 9/

VOID: REI 12 03/26/24 13:43

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Monroe 104

Operator Name: I.D.

Wrisant City PD

Location

AS 1U Serial no: 035714
Version no: 532B

TEST RECORD 01558

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:28 .000

Calibration Check: 24 03/26/24 13:28 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Operator Name: I.D. Monroe 104 230040

Location Wright City PD

AS 1U Serial no: 035714
Version no: 532B

TEST RECORD 01559

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:34 .000

Calibration Check: 24 03/26/24 13:34 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Operator Name: I.D. Monroe 104 230040

Location Wright City PD

AS 1U Serial no: 035714
Version no: 532B

TEST RECORD 01560

Temp Date Time 210L 9/

Air Blank: 03/26/24 13:38 .000

Calibration Check: 25 03/26/24 13:38 .102

Monroe 104

Subject Name

Monroe 104

Subject I.D.

Operator Name: I.D. Monroe 104 230040

Location Wright City PD

AS 1U Serial no: 035714
Version no: 532B

TEST RECORD 01561

Temp Date Time 210L 9/

VOID: REI 12 03/26/24 13:43

Monroe 104


Subject Name

Monroe 104

Subject I.D.

Operator Name: I.D. Monroe 104

Location Wright City PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

GREGORY D. MONROE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo. and 306.111 through 306.119 RSMo.

DATE: 3/27/24

NUMBER: 230120

EXPIRES: 3/27/25

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Daniel J. McKeel

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PB-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

**PERMIT
TYPE III**

JOSHUA A. KRULL

is hereby authorized to operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo. and 306.111 through 306.119 RSMo.