



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087971	NAME OF AGENCY CARROLLTON POLICE DEPARTMENT	DATE OF INSPECTION 04/04/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 1320 US 65 N., CARROLLTON, MO, 64633		TIME OF INSPECTION 4:46 A.M.

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LABS INC. LOT # 22430 EXP. DATE 11/30/2024

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2940 SIM. NIST EXP DATE 09/19/2024

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .103	TEST 2  .103	TEST 3  .102
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Update Time Due To Daylight Savings  
 Maintenance conducted in accordance with DHSS standards.

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME WILLIAM BAGER
TYPE II PERMIT NUMBER/EXPIRATION DATE 230029/2-17-2025	TELEPHONE NUMBER (660) 542-3128

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

CARROLLTON POLICE DEPARTMENT-BREATH ALCOHOL PROGRAM

AS IV Serial no: 087971  
Version no: 5028

TEST RECORD - RERRINT

TEST RECORD 00501

Temp Date Time 210L 9/

Air Blank:  
04/04/24 04:45 .002  
Calibration Check:  
29 04/04/24 04:46 .103

Subject Name  
TEST #1  
Subject I.D.

AS IV Serial no: 087971  
Version no: 5028

TEST RECORD 00502

Temp Date Time 210L 9/

Air Blank:  
04/04/24 04:48 .003  
Calibration Check:  
24 04/04/24 04:48 .103

Subject Name  
TEST #2  
Subject I.D.

AS IV Serial no: 087971  
Version no: 5028

TEST RECORD 00503

Temp Date Time 210L 9/

Air Blank:  
04/04/24 04:50 .002  
Calibration Check:  
24 04/04/24 04:50 .102

Subject Name  
TEST #3  
Subject I.D.

AS IV Serial no: 087971  
Version no: 5028

TEST RECORD 00504

Temp Date Time 210L 9/

VOID: RTI  
12 04/04/24 04:52

Subject Name  
RFI TEST  
Subject I.D.

Operator Name: I.D.

W. BARGER #230029

Location

1320 N US 65

CARROLLTON, MO

Operator Name: I.D.

W. BARGER #230029

Location

1320 N US 65

CARROLLTON, MO

Operator Name: I.D.

W. BARGER #230029

Location

1320 N US 65

CARROLLTON, MO

Operator Name: I.D.

W. BARGER #230029

Location

1320 N US 65

CARROLLTON, MO

TEST #1

TEST #2

TEST #3

RFI TEST

AS-IV S/N: 087971 Type II Permit Holder: William Barger #230029-Exp: 02/17/2025 Date of Maintenance Report: 4/4/2024

Carrollton Police Department AS-IV Monthly Maintenance Report



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-684-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**WILLIAM BARGER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/17/2023

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230029

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 2/17/2025

MO 680-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BARGER, WILLIAM  
 Permit No 230029  
 Date Issued 2/17/2023 Date Expires 2/17/2025

