

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 01089
Temp Date Time 210L
Air Blank: 01/26/24 18:17 .000
Calibration Check: 15 01/26/24 18:17 .098

Subject Name
TEST 1
Subject I.D.

Operator Name, I.D.
SIMPSON 230135

Location
1000 CITY PKWY
OSAGE BEACH, MO

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 01090
Temp Date Time 210L
Air Blank: 01/26/24 18:21 .000
Calibration Check: 16 01/26/24 18:21 .098

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
SIMPSON, 230135

Location
1000 CITY PKWY,
OSAGE BEACH, MO

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 01091
Temp Date Time 210L
Air Blank: 01/26/24 18:25 .000
Calibration Check: 17 01/26/24 18:25 .098

Subject Name
TEST 3
Subject I.D.

Operator Name, I.D.
SIMPSON 230135

Location
1000 CITY PARKWAY
OSAGE BEACH, MO

AS IV Serial no: 107995
Version no: 532B

TEST RECORD - REPRINT
TEST RECORD 01092
Temp Date Time 210L
VOID: RFI
12 01/26/24 18:28

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
SIMPSON 230135

Location
1000 CITY PKWY
OSAGE BEACH, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN1172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JUSTICE SIMPSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/6/2023

NUMBER 230135

EXPIRES 7/6/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nickelsson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SIMPSON, JUSTICE
 Permit No 230135
 Date Issued 7/6/2023 Date Expires 7/6/2025

