



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 8:27 am, Feb 02, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107999	NAME OF AGENCY Clay County Sheriff's Office	DATE OF INSPECTION 02/01/2024
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 12 S. Water St., Liberty, MO 64068	TIME OF INSPECTION 2025
--	----------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG221502</u> EXP. DATE <u>08/03/2024</u>	
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE	

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .100	TEST 3 .099
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
 Repaired at Safety Center. Placed back into service.

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Tom Butkovich
TYPE II PERMIT NUMBER/EXPIRATION DATE 230113 - 06/05/2025	TELEPHONE NUMBER 8164073702

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00453

Temp Date Time 210L
9/

Air Blank:
02/01/24 20:30 .000
Calibration Check:
19 02/01/24 20:30 .101

Subject Name

Test #1

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00454

Temp Date Time 210L
9/

Air Blank:
02/01/24 20:32 .000
Calibration Check:
19 02/01/24 20:32 .100

Subject Name

Test # 2

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00455

Temp Date Time 210L
9/

Air Blank:
02/01/24 20:34 .000
Calibration Check:
21 02/01/24 20:34 .099

Subject Name

Test #3

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00456

Temp Date Time 210L
9/

VOID: REF
12 02/01/24 20:36

Subject Name

REF Test

Subject I.D.

Butkovich 2260

Operator Name, I.D.

CCSO

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 27-Jan-2021

Lot # AG102503 Model 108caccd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
25-Jan-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
THOMAS F. BUTKOVICH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/5/2023

NUMBER 230113

EXPIRES 6/5/2025

Mike Masoma

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BUTKOVICH, THOMAS
 Permit No 230113
 Date Issued 6/5/2023 Date Expires 6/5/2025

