



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 7:05 am, Mar 01, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111328	NAME OF AGENCY Knob Noster PD	DATE OF INSPECTION 02/29/2024
-----------------------------	----------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 201 N. State Street, Knob Noster	TIME OF INSPECTION 4:02 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>22430</u> EXP. DATE <u>11/30/2024</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u> SIM. SN <u>SD2231</u> SIM. NIST EXP DATE <u>11/16/2024</u>	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101	TEST 2 ← .101	TEST 3 ← .100
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).  
 Instrument working correctly within Dept. of Health Standards.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Karl Van Vickle
TYPE II PERMIT NUMBER/EXPIRATION DATE 220152 / 05-25-2024	TELEPHONE NUMBER (660) 563-2233

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111328  
Version no: 532B

TEST RECORD 00799

Temp Date Time 210L

Air Blank:  
02/29/24 16:02 .000  
Calibration Check:  
22 02/29/24 16:02 .000  
Monthly Maint  
Subject Name

Blank Standard  
Subject I.D.

KivanVickle 220152  
Operator Name, I.D.

201 N. State St.  
Location

Knob Noster, MO 65336

Knob Noster PD

AS IV Serial no: 111328  
Version no: 532B

TEST RECORD 00800

Temp Date Time 210L

Air Blank:  
02/29/24 16:03 .000  
Calibration Check:  
22 02/29/24 16:03 .101  
TEST #1 Monthly  
Subject Name Maint

Test #1  
KivanVickle 220152  
Subject I.D.

KivanVickle 220152  
Operator Name, I.D.

201 N. State St.  
Location

Knob Noster MO 65336

Knob Noster PD

AS IV Serial no: 111328  
Version no: 532B

TEST RECORD , 00801

Temp Date Time 210L

Air Blank:  
02/29/24 16:05 .000  
Calibration Check:  
22 02/29/24 16:05 .101  
Monthly Maint  
Subject Name

Test #2  
Subject I.D.

KivanVickle 220152  
Operator Name, I.D.

201 N. State St.  
Location

Knob Noster MO 65336

Knob Noster PD

AS IV Serial no: 111328  
Version no: 532B

TEST RECORD 00802

Temp Date Time 210L

Air Blank:  
02/29/24 16:06 .000  
Calibration Check:  
23 02/29/24 16:06 .100  
Monthly Maint  
Subject Name

Test #3  
Subject I.D.

KivanVickle 220152  
Operator Name, I.D.

201 N. State St.  
Location

Knob Noster MO 65336

Knob Noster PD

AS IV Serial no: 111328  
Version no: 532B

TEST RECORD 00803

Temp Date Time 210L

VOID: RFI  
12 02/29/24 16:08  
Monthly Maint  
Subject Name

RFI CHECK  
Subject I.D.

KivanVickle 220152  
Operator Name, I.D.

201 N. State St.  
Location

Knob Noster MO 65336

Knob Noster PD



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET ◦ HARRISBURG, PA 17111-4511 ◦ TELEPHONE: 717-864-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 22430 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on December 1, 2022, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is November 30, 2024 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

COPY



Paula Nickelson  
 Acting Director

Michael L. Parson  
 Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

Simulator Serial Number: SD2231      Manufacturer: Guth  
 Model Number: 10-4D  
 Agency: KNOB NOSTER PD  
 Agency Address: 201 N STATE, KNOB NOSTER, MO 65336

## NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689      Bias: 0.00  
 Uncertainty: 0.02  
 Date of Certification: 1/23/2023      Date of Expiration: 1/23/2024

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.99	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 11/16/2023  
 Certification Expiration: 11/16/2024  
 Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: SD2231\_11162023

X *Brianna Medrano*

DHSS BAP Scientist Approving

COP



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

PERMIT  
 TYPE II

KARL E. VANVICKLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 208.020 through 208.04-1, RSMo and 306.111 through 306.119 RSMo.

ISSUED 5/25/2022  
 PERMIT NUMBER 220152  
 EXPIRES 5/25/2024

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4-(R6-10)

0771 (6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate any evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator VANVICKLE, KARL  
 Permit No 220152  
 Date Issued 5/25/2022 Date Expires 5/25/2024



COPY