

AS IV Serial no: 111656
Version no: 532B

TEST RECORD 00094

Temp Date Time 210L
s/

Air Blank:
04/04/24 13:26 .000
Calibration Check:
24 04/04/24 13:26 .102

Subject Name

Test

Subject I.D.

Test 1

Operator Name, I.D.

William Housley 230168

Location

517 Onada St.

Seneca Mo 64865

AS IV Serial no: 111656
Version no: 532B

TEST RECORD 00095

Temp Date Time 210L
s/

Air Blank:
04/04/24 13:27 .000
Calibration Check:
24 04/04/24 13:27 .101

Subject Name

Test

Subject I.D.

Test 2

Operator Name, I.D.

William Housley 230168

Location

517 Onada St.

Seneca Mo 64865

AS IV Serial no: 111656
Version no: 532B

TEST RECORD 00096

Temp Date Time 210L
s/

Air Blank:
04/04/24 13:29 .000
Calibration Check:
25 04/04/24 13:29 .101

Subject Name

Test

Subject I.D.

Test 3

Operator Name, I.D.

William Housley, AT 230168

Location

517 Onada St.

Seneca Mo 64865

AS IV Serial no: 111656
Version no: 532B

TEST RECORD 00097

Temp Date Time 210L
s/

VOID: RFI
12 04/04/24 13:30

Subject Name

Test

Subject I.D.

RFI Test.


Operator Name, I.D.

William Housley 230168

Location

517 Onada St.

Seneca Mo 64865



GUTH LABORATORIES, INC.

580 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23390 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1207% (w/vol) ethyl alcohol. The expiration date for this lot number is October 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- 0.2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

**PERMIT
TYPE II**

WILLIAM J. HOUSLEY II

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs,
and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 8/7/2023
NUMBER 230168
EXPIRES 8/7/2025

MO 680-0774 (8-10)


Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul J. Nicholas

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (88-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **HOUSLEY II, WILLIAM**
Permit No **230168**
Date Issued **8/7/2023** Date Expires **8/7/2025**

