



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 2:09 pm, Feb 05, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111674	NAME OF AGENCY MSHP	DATE OF INSPECTION 02/02/2024
-----------------------------	------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 102 W N Main St, Richmond, Missouri 64085	TIME OF INSPECTION 6:10 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u>	LOT # <u>23390</u> EXP. DATE <u>10/17/2025</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.00</u>	SIM. SN <u>MP2125</u> SIM. NIST EXP DATE <u>11/27/2024</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100	TEST 2 .100	TEST 3 .102
--------------	--------------	--------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09) 2	(.10-.14) 3	(.15-.19)	(OVER .19)
----------	---------	--------------------	--------------------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Isaac Kimball
TYPE II PERMIT NUMBER/EXPIRATION DATE 230322 12/21/2025	TELEPHONE NUMBER (816) 622-0800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00596

Temp	Date	Time	a/ 210L
Air Blank:			
	02/02/24	18:10	.000
Calibration Check:			
	19 02/02/24	18:10	.100

Subject Name
Test

Subject I.D.
1

Operator Name, I.D.
1

Operator Name, I.D.
1

Location
Tr. I. Kimball #883

Location
Zone 5 office

Location
Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00597

Temp	Date	Time	a/ 210L
Air Blank:			
	02/02/24	18:12	.000
Calibration Check:			
	20 02/02/24	18:12	.100

Subject Name
Test

Subject I.D.
2

Operator Name, I.D.
2

Operator Name, I.D.
2

Operator Name, I.D.
2

Location
Tr. I. Kimball #883

Location
Zone 5 office

Location
Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00598

Temp	Date	Time	a/ 210L
Air Blank:			
	02/02/24	18:14	.000
Calibration Check:			
	21 02/02/24	18:14	.102

Subject Name
Test

Subject I.D.
3

Operator Name, I.D.
3

Operator Name, I.D.
3

Operator Name, I.D.
3

Location
Tr. I. Kimball #883

Location
Zone 5 office

Location
Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00599

Temp	Date	Time	a/ 210L
VOID: RFI			
	12 02/02/24	18:16	

Subject Name
RFI

Subject I.D.
RFI

Operator Name, I.D.
RFI

Operator Name, I.D.
RFI

Operator Name, I.D.
RFI

Location
Tr. I. Kimball #883

Location
Zone 5 office

Location
Zone 5 office



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ISAAC KIMBALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

NUMBER 230322

EXPIRES 12/21/2025

Mike Masman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KIMBALL, ISAAC
Permit No 230322
Date Issued 12/21/2023 **Date Expires** 12/21/2025

