

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00600

Temp Date Time ^{a/} 210L

Air Blank:
03/03/24 20:26 .000

Calibration Check:
21 03/03/24 20:26 .102

Subject Name

Test

Subject I.D.

1

Operator Name, I.D.

Tr. I. Kimball #883

Location

Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00601

Temp Date Time ^{a/} 210L

Air Blank:
03/03/24 20:28 .000

Calibration Check:
23 03/03/24 20:28 .102

Subject Name

Test

Subject I.D.

2

Operator Name, I.D.

Tr. I. Kimball #883

Location

Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00602

Temp Date Time ^{a/} 210L

Air Blank:
03/03/24 20:29 .000

Calibration Check:
23 03/03/24 20:29 .102

Subject Name

Test

Subject I.D.

3

Operator Name, I.D.

Tr. I. Kimball #883

Location

Zone 5 office

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00603

Temp Date Time ^{a/} 210L

VOID: RFI
12 03/03/24 20:31

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Tr. I. Kimball #883

Location

Zone 5 office



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ISAAC KIMBALL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230322

Dave J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 12/21/2025

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KIMBALL, ISAAC
 Permit No 230322
 Date Issued 12/21/2023 Date Expires 12/21/2025

