



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 7:16 am, Jan 02, 2024

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111677	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 01/01/2024
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) MADISON COUNTY JAIL (124 N. MAIN, FREDERICKTOWN, MO 63645)	TIME OF INSPECTION 4:30 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABROATORIES LOT # 22430 EXP. DATE 11/30/2024
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIM. SN MP2207 SIM. NIST EXP DATE 03/01/2024

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .099	TEST 3 .099
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME ADAM SHIPLEY
TYPE II PERMIT NUMBER/EXPIRATION DATE 230275 / 11/28/2025	TELEPHONE NUMBER (573) 840-9500

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 01107

Temp Date Time ^{s/} 210L

Air Blank:
01/01/24 16:37 .000
Calibration Check:
22 01/01/24 16:37 .099

Subject Name

Test 1
Subject I.D.

Operator Name, I.D.

AR Shipley 230275
Location
Madison Co Jail

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 01108

Temp Date Time ^{s/} 210L

Air Blank:
01/01/24 16:40 .000
Calibration Check:
22 01/01/24 16:40 .099

Subject Name

Test 2
Subject I.D.

Operator Name, I.D.

AR Shipley 230275
Location
Madison Co Jail

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 01109

Temp Date Time ^{s/} 210L

Air Blank:
01/01/24 16:42 .000
Calibration Check:
23 01/01/24 16:42 .099

Subject Name

Test 3
Subject I.D.

Operator Name, I.D.

AR Shipley 230275
Location
Madison Co Jail

AS IV Serial no: 111677
Version no: 532B

TEST RECORD 01110

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/01/24 16:46

Subject Name

RFI Test
Subject I.D.

Operator Name, I.D.

AR Shipley 230275
Location
Madison Co Jail



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ADAM SHIPLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/28/2023

NUMBER 230275

EXPIRES 11/28/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SHIPLEY, ADAM
Permit No 230275
Date Issued 11/28/2023 **Date Expires** 11/28/2025

