



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111737 PRINTER #09B.3589.513	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 01/23/2024
---	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 3525 NORTH BELT HIGHWAY, ST. JOSEPH, MO, 64506	TIME OF INSPECTION 11:00 am
--	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>GUTH</u>	LOT # <u>23180</u> EXP. DATE <u>05/17/2025</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u>	SIM. SN <u>MP2502</u> SIM. NIST EXP DATE <u>02/06/2024</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .103	TEST 2 • .102	TEST 3 • .102
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME MICHAEL J. MILLER
TYPE II PERMIT NUMBER/EXPIRATION DATE 202258 11/17/2024	TELEPHONE NUMBER (816) 387-2345

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111737  
Version no: 532B

TEST RECORD 00108

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:  
01/23/24 11:05 .000  
Calibration Check:  
23 01/23/24 11:05 .103

Subject Name

MAINTENANCE #1

Subject I.D.

N/A

Operator Name, I.D.

M.J. MILLER #1418

Location

3525 N BELT HWY

ST. JOSEPH, MO, 64506

AS IV Serial no: 111737  
Version no: 532B

TEST RECORD 00109

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:  
01/23/24 11:09 .000  
Calibration Check:  
24 01/23/24 11:09 .102

Subject Name

SELF TEST #2

Subject I.D.

N/A

Operator Name, I.D.

M.J. MILLER #1418

Location

3525 N BELT HWY

ST. JOSEPH, MO, 64506

TEST RECORD 00110

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:  
01/23/24 11:11 .000  
Calibration Check:  
25 01/23/24 11:11 .102

Subject Name

MAINTENANCE #3

Subject I.D.

N/A

Operator Name, I.D.

M.J. MILLER #1418

Location

3525 N BELT HWY

ST. JOSEPH, MO, 64506

AS IV Serial no: 111737  
Version no: 532B

TEST RECORD 00111

Temp	Date	Time	a/ 210L
------	------	------	------------

VOID: RFI  
12 01/23/24 11:13

Subject Name

MAINTENANCE RFI

Subject I.D.

N/A

Operator Name, I.D.

M.J. MILLER #1418

Location

3525 N BELT HWY

ST. JOSEPH, MO, 64506

M.J. MILLER #1418



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23180** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1220%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN1172002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MICHAEL J. MILLER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/17/2022

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220258

*Douglas F. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 11/17/2024

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator MILLER, MICHAEL  
 Permit No 220258  
 Date Issued 11/17/2022 Date Expires 11/17/2024

