



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111744	NAME OF AGENCY BOONE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 03/14/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E. COUNTY DRIVE, COLUMBIA, MISSOURI 65202		TIME OF INSPECTION 11:50 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG306503 EXP. DATE 03/06/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .102

TEST 3 .102

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Due to daylight savings time, a time discrepancy of one hour was corrected, otherwise the instrument was operating per the Department of Health and Senior Services standards.

**INSPECTING OFFICER**

SIGNATURE *Kevin Stroer*

PRINT NAME  
KEVIN STROER

TYPE II PERMIT NUMBER/EXPIRATION DATE  
#220139 / EXPIRES 05/11/2024

TELEPHONE NUMBER  
(573) 875-1111

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IU Serial no: 111744  
Version no: 532B

TEST RECORD 00819

Temp Date Time 210L  
g/

Air Blank: 03/14/24 23:50 .000  
Calibration Check: 23 03/14/24 23:50 .102

Subject Name

Monthly Maintenance

Subject I.D.

Test # 1

Operator Name: I.D.  
Kevin J. Strober

Permit # 220139

Location

2111 E. County Drive,

Columbia, MD 05202

AS IU Serial no: 111744  
Version no: 532B

TEST RECORD 00820

Temp Date Time 210L  
g/

Air Blank: 03/14/24 23:54 .000  
Calibration Check: 24 03/14/24 23:54 .102

Subject Name

Monthly Maintenance

Subject I.D.

Test # 2

Operator Name: I.D.  
Kevin J. Strober

Permit # 220139

Location

2111 E. County Drive,

Columbia, MD 05202

AS IU Serial no: 111744  
Version no: 532B

TEST RECORD 00821

Temp Date Time 210L  
g/

Air Blank: 03/14/24 23:59 .000  
Calibration Check: 24 03/14/24 23:59 .102

Subject Name

Monthly Maintenance

Subject I.D.

Test # 3

Operator Name: I.D.  
Kevin J. Strober

Permit # 220139

Location

2111 E. County Drive,

Columbia, MD 05202

AS IU Serial no: 111744  
Version no: 532B

TEST RECORD 00822

Temp Date Time 210L  
g/

VOID: RFI  
12 03/15/24 00:02

Subject Name

Monthly Maintenance

Subject I.D.

Test # 4; VOID RFI

Operator Name: I.D.  
Kevin J. Strober

Permit # 220139

Location

2111 E. County Drive,

Columbia, MD 05202



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 7-Mar-2023

**Lot #** AG306503 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
6-Mar-2025	108	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
Reason:Dry gas standard certification of analysis  
Location:Airgas USA LLC (Lab)  
Date:03.09.2023 20:47

Approved for Release: \_\_\_\_\_

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**KEVIN STROER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220139

EXPIRES 5/11/2024

*Laura P. Way*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nicholson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

**STATE OF MISSOURI**  
**DEPARTMENT OF HEALTH AND SENIOR SERVICES**  
**BREATH ALCOHOL PROGRAM**

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** STROER, KEVIN  
**Permit No** 220139  
**Date Issued** 5/11/2022 **Date Expires** 5/11/2024

