



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111751	NAME OF AGENCY BOONE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 03/15/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E. COUNTY DRIVE, COLUMBIA, MISSOURI 65202		TIME OF INSPECTION 0:44 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER INTOXIMETERS LOT # AG306503 EXP. DATE 03/06/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103	TEST 2 .101	TEST 3 .101
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Due to daylight savings time, a time discrepancy of one hour was corrected, otherwise the instrument was operating per the Department of Health and Senior Services standards.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME KEVIN STROER
TYPE II PERMIT NUMBER/EXPIRATION DATE #220139 / EXPIRES 05/11/2024	TELEPHONE NUMBER (573) 875-1111

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00766

Temp Date Time 210L
s/

Air Blank: 03/15/24 00:44 .000
Calibration Check: 22 03/15/24 00:44 .103

Subject Name

Monthly Maintenance

Subject I.D.

Test #1

Operator Name, I.D.
Kevin J. Stroer

Permit # 220139

Location

2111 E. County Drive,

Columbia, MO 65202

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00767

Temp Date Time 210L
s/

Air Blank: 03/15/24 00:50 .000
Calibration Check: 23 03/15/24 00:50 .101

Subject Name

Monthly Maintenance

Subject I.D.

Test #2

Operator Name, I.D.
Kevin J. Stroer

Permit # 220139

Location

2111 E. County Drive

Columbia, MO 65202

AS IV Serial no: 111751
Version no: 532B

TEST RECORD 00768

Temp Date Time 210L
s/

Air Blank: 03/15/24 00:56 .000
Calibration Check: 23 03/15/24 00:56 .101

Subject Name

Monthly Maintenance

Subject I.D.

Test #3

Operator Name, I.D.
Kevin J. Stroer

Permit # 220139

Location

2111 E. County Drive,

Columbia, MO 65202

AS IV Serial no: 111751
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00769

Temp Date Time 210L
s/

Air Blank: 08/19/10 14:45 0.000
Calibration1: 10 08/19/10 14:45 0.029

Subject Name

Monthly Maintenance

Subject I.D.

Test #4, VOID REI

Operator Name, I.D.
Kevin J. Stroer

Permit # 220139

Location

2111 E. County Drive,

Columbia, MO 65202



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Mar-2023

Lot # AG306503 **Model** 108

Exp Date 6-Mar-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (260 ppm)
-------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:03.09.2023 20:47

Approved for Release: 
 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KEVIN STROER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220139

EXPIRES 5/11/2024

Laura P. Way
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielsen
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STROER, KEVIN
Permit No 220139
Date Issued 5/11/2022 **Date Expires** 5/11/2024

