

AS IV Serial no: 111769
Version no: 532B

TEST RECORD 00456

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 19:11 .000
Calibration Check:
27 12/28/23 19:11 .100

Subject Name

Test # 1

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 111769
Version no: 532B

TEST RECORD 00457

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 19:13 .000
Calibration Check:
27 12/28/23 19:13 .100

Subject Name

Test # 2

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 111769
Version no: 532B

TEST RECORD 00458

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 19:14 .000
Calibration Check:
27 12/28/23 19:14 .101

Subject Name

Test # 3

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 111769
Version no: 532B

TEST RECORD 00459

Temp Date Time ^{g/} 210L

VOID: RFI
12 12/28/23 19:16

Subject Name

RFI check

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 111769
Version no: 532B

TEST RECORD 00460

Temp Date Time ^{g/} 210L

Air Blank:
12/28/23 19:17 .000
Subject Test: Auto
26 12/28/23 19:17 .000

Subject Name

Self Test

Subject I.D.

Hochne, Mark D. 2078
Operator Name, I.D.

Location



Airgas USA LLC (LAB)
 3600 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-Apr-2022

Lot # AG209701 Model 108

Exp Date 7-Apr-2024	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
------------------------	------------------	----------------------------------	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by: Quality Control
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)
 Date: 04.12.2022 16:30

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MARK D HOEHNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/14/2022

NUMBER 220158

EXPIRES 6/14/2024

Mike Magama

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David T. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-9771 (8-10)

LAD-4 (RB-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an optional breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **HOEHNE, MARK**
Permit No **220158**
Date Issued **6/14/2022** Date Expires **6/14/2024**

