



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 113777	NAME OF AGENCY St. Louis County Police Department	DATE OF INSPECTION 03/29/2024
LOCATION OF INSTRUMENT (STREET AND CITY) 14301 SOUTH OUTER FOURTY MODOT TMC		TIME OF INSPECTION 11:20 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG321505 EXP. DATE 08/03/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .100	TEST 2 → .100	TEST 3 → .099
---------------	---------------	---------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME PO Tosie DSN 4553
TYPE #/PERMIT NUMBER/EXPIRATION DATE 220173 / 06242024	TELEPHONE NUMBER (636) 529-8210

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 113777  
Version no: 532C

TEST RECORD 00364

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
03/29/24 11:20 .000  
Calibration Check:  
18 03/29/24 11:20 .100

Subject Name

Test #1  
Subject I.D.

N/A  
Operator Name, I.D.

P.O. Tosre 4553  
Location

MOBOT TMC

AS IV Serial no: 113777  
Version no: 532C

TEST RECORD 00365

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
03/29/24 11:22 .000  
Calibration Check:  
18 03/29/24 11:22 .100

Subject Name

Test #2  
Subject I.D.

N/A  
Operator Name, I.D.

P.O. Tosre 4553  
Location

MOBOT TMC

AS IV Serial no: 113777  
Version no: 532C

TEST RECORD 00366

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
03/29/24 11:23 .000  
Calibration Check:  
19 03/29/24 11:23 .099

Subject Name

Test #3  
Subject I.D.

N/A  
Operator Name, I.D.

P.O. Tosre 4553  
Location

MOBOT TMC

AS IV Serial no: 113777  
Version no: 532C

TEST RECORD 00367

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 03/29/24 11:25

Subject Name

RFI  
Subject I.D.

N/A  
Operator Name, I.D.

P.O. Tosre 4553  
Location

MOBOT TMC



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
*Exclusive Supplier*  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 3-Aug-2023

**Lot # AG321505 Model 108**

<b>Exp Date</b> 3-Aug-2025	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
-------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

**Analytical Method:** NDIR

Digitally signed by: Quality Control  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)  
 Date: 08.03.2023 17:58



**Approved for Release:** \_\_\_\_\_  
 Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**ROBERT TOSIE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220173

EXPIRES 6/24/2024

*Mike Morrison*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Dave F. Richardson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES