



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------|---|----------------------------------|
| ALCO SENSOR IV SN 119737 | NAME OF AGENCY MARYVILLE POLICE DEPARTMENT | DATE OF INSPECTION 01/29/2024 |
|-----------------------------|---|----------------------------------|

| | |
|--|-------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 101 N. VINE ST., MARYVILLE, MO 64468 | TIME OF INSPECTION 8:45 pm |
|--|-------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|--|--|
| <input checked="" type="checkbox"/> SIMULATOR SOLUTION | <input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>GUTH</u> | LOT # <u>22430</u> EXP. DATE <u>11/30/2024</u> |
| <input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34</u> | SIM. SN <u>SD2281</u> SIM. NIST EXP DATE <u>08/18/2024</u> |

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|--------------|--------------|--------------|
| TEST 1 .102 | TEST 2 .102 | TEST 3 .100 |
|--------------|--------------|--------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT NAME WAYNE L. WILSON |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220286 / 12/21/2024 | TELEPHONE NUMBER (660) 562-3209 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119737
Version no: 532C

TEST RECORD 00141

Temp Date Time ^{g/} 210L

Air Blank:
01/29/24 20:42 .000
Calibration Check:
21 01/29/24 20:42 .102

Subject Name
Test #1

Subject I.D.

Operator Name, I.D.
Wayne Wilson/220286

Location
Maryville PD

101 N. Vine, Maryville
MO

AS IV Serial no: 119737
Version no: 532C

TEST RECORD 00142

Temp Date Time ^{g/} 210L

Air Blank:
01/29/24 20:43 .000
Calibration Check:
22 01/29/24 20:43 .102

Subject Name
Test #2

Subject I.D.

Operator Name, I.D.
Wayne Wilson/220286

Location
Maryville PD

101 N. Vine, Maryville
MO

AS IV Serial no: 119737
Version no: 532C

TEST RECORD 00143

Temp Date Time ^{g/} 210L

Air Blank:
01/29/24 20:45 .000
Calibration Check:
23 01/29/24 20:45 .100

Subject Name
Test #3

Subject I.D.

Operator Name, I.D.
Wayne Wilson/220286

Location
Maryville PD

101 N. Vine, Maryville
MO

AS IV Serial no: 119737
Version no: 532C

TEST RECORD 00144

Temp Date Time ^{g/} 210L

VOID: RFI
12 01/29/24 20:47

Subject Name
RFI Test

Subject I.D.

Operator Name, I.D.
Wayne Wilson/220286

Location
Maryville PD

101 N. Vine, Maryville,
MO

AS IV Serial no: 119737
Version no: 532C

TEST RECORD 00145

Temp Date Time ^{g/} 210L

Air Blank:
01/29/24 20:48 .000
Subject Test: Auto
23 01/29/24 20:48 .000

Subject Name
Self Test

Subject I.D.

Operator Name, I.D.
Wayne Wilson/220286

Location
Maryville PD

101 N. Vine, Maryville
MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

ORIGINAL

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22430** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 1, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 30, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

ORIGINAL



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
WAYNE L. WILSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2022

NUMBER 220286

EXPIRES 12/21/2024

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Doreen J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-6771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WILSON, WAYNE
Permit No 220286
Date Issued 12/21/2022 **Date Expires** 12/21/2024

