



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

REPORT #

By Tracy Crews at 7:20 am, Apr 12, 2024

Complete this report in duplicate at the time of the regular monthly preventative maintenance check. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <i>120501</i>	NAME OF AGENCY DOI-National Park Service	DATE OF INSPECTION <i>4/11/24</i>
------------------------------------	---	--------------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Akers Ranger Station-Intersection of Highways K and KK	TIME OF INSPECTION <i>1444</i>
--	-----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 23180 EXP. DATE 05/17/2025
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP7615 SIM. NIST EXP DATE 10/18/2024
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <i>.100</i>	TEST 2 <i>.100</i>	TEST 3 <i>.099</i>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <i>Zach Swindle</i>	PRINT NAME Zachary Swindle
TYPE II PERMIT NUMBER/EXPIRATION DATE 230181 08/17/2025	TELEPHONE NUMBER (573) 351-9645

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 120501
Version no: 532C

TEST RECORD 00038 g/

Temp Date Time 210L
VOID: RFI
12 04/11/24 14:48

Subject Name
Jack Swindle
Subject I.D.
230181 8/17/25
Operator Name, I.D.
AKERS RANGER STATION
Location
HWY K & RK
INTERSECTION

AS IV Serial no: 120501
Version no: 532C

TEST RECORD 00037 g/

Temp Date Time 210L
Air Blank: 04/11/24 14:47 .000
Calibration Check: 22 04/11/24 14:47 .099

Subject Name
Jack Swindle
Subject I.D.
230181 8/17/25
Operator Name, I.D.
AKERS RANGER STATION
Location
HWY K & RK
INTERSECTION

AS IV Serial no: 120501
Version no: 532C

TEST RECORD 00036 g/

Temp Date Time 210L
Air Blank: 04/11/24 14:46 .000
Calibration Check: 21 04/11/24 14:46 .100

Subject Name
Jack Swindle
Subject I.D.
230181 8/17/25
Operator Name, I.D.
AKERS RANGER STATION
Location
HWY K & RK
INTERSECTION

AS IV Serial no: 120501
Version no: 532C

TEST RECORD 00035 g/

Temp Date Time 210L
Air Blank: 04/11/24 14:44 .000
Calibration Check: 21 04/11/24 14:44 .100

Subject Name
Jack Swindle
Subject I.D.
230181 8/17/25
Operator Name, I.D.
AKERS RANGER STATION
Location
HWY K & RK
INTERSECTION



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23180 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1220% (w/vol) ethyl alcohol. The expiration date for this lot number is May 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN11172002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP7615 Manufacturer: Guth
Model Number: 12V500
Agency: NATIONAL PARK SERVICE
Agency Address: 404 WATERCROSS, VAN BUREN, MO 63965

NIST THERMOMETER INFORMATION

Serial Number: 19BMM01308 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 4/20/2023 Date of Expiration: 4/20/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.00), Combined Uncertainty (.02)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 10/18/2023
Certification Expiration: 10/18/2024
Simulator testing technician: B. LUTMER

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: MP7615_10182023

X [Signature]

DHSS BAP Scientist Approving



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ZACHARY SWINDLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/17/2023

Mike Masoma

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230181

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 8/17/2025

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SWINDLE, ZACHARY
Permit No 230181
Date Issued 8/17/2023 **Date Expires** 8/17/2025

