



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12679	NAME OF AGENCY Chesterfield PD	DATE OF INSPECTION 02/27/2024																																																				
LOCATION OF INSTRUMENT (STREET AND CITY) 690 Chesterfield Pkwy W Chesterfield		TIME OF INSPECTION 14:52 CST																																																				
<p>CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.</p> <p><input checked="" type="checkbox"/> DIAGNOSTIC RECORD</p> <table border="0"> <tr> <td><input checked="" type="checkbox"/> BLANK CHECK</td> <td><input checked="" type="checkbox"/> CO2 CHECK</td> </tr> <tr> <td><input checked="" type="checkbox"/> FC 1 TEMP</td> <td><input checked="" type="checkbox"/> FLOW CHECK</td> </tr> <tr> <td><input checked="" type="checkbox"/> SRC TEMP</td> <td><input checked="" type="checkbox"/> FCB CHECK</td> </tr> <tr> <td><input checked="" type="checkbox"/> DET TEMP</td> <td><input checked="" type="checkbox"/> CRC COMP CHECK</td> </tr> <tr> <td><input checked="" type="checkbox"/> BT TEMP</td> <td><input checked="" type="checkbox"/> CRC CAL CHECK</td> </tr> <tr> <td><input checked="" type="checkbox"/> STD 2 TEMP</td> <td><input checked="" type="checkbox"/> PRINT TEST</td> </tr> <tr> <td><input checked="" type="checkbox"/> ETH CHECK</td> <td></td> </tr> </table> <p>BREATH ANALYZER ACCURACY STANDARDS</p> <table border="0"> <tr> <td><input type="checkbox"/> SIMULATOR SOLUTION</td> <td><input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE</td> </tr> <tr> <td><input checked="" type="checkbox"/> STANDARD SUPPLIER</td> <td>INTOXIMETERS</td> </tr> <tr> <td></td> <td>LOT# AG223501</td> </tr> <tr> <td></td> <td>EXP. DATE 08/23/2024</td> </tr> <tr> <td><input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)</td> <td>SIM. SN</td> </tr> <tr> <td></td> <td>SIM. NIST EXP DATE</td> </tr> </table> <p><input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)</p> <p>Run three tests using a standard solution. All three tests must be within <u>+5%</u> of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.</p> <table border="0"> <tr> <td><input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE</td> </tr> <tr> <td><input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE</td> </tr> <tr> <td><input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE</td> </tr> </table> <table border="1"> <tr> <td>TEST 1 @ 0.079 g/210L</td> <td>TEST 2 @ 0.079 g/210L</td> <td>TEST 3 @ 0.079 g/210L</td> </tr> </table> <p>INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:</p> <table border="1"> <tr> <td>REFUSALS</td> <td>1</td> <td>0-.04</td> <td>11</td> <td>.05-.09</td> <td>0</td> <td>.10-.14</td> <td>1</td> <td>.15-.19</td> <td>1</td> <td>OVER .19</td> <td>1</td> </tr> </table> <p>LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).</p> <p>INSTRUMENT IS OPERATING WITHIN MO DEPT OF HEALTH SPECIFICATIONS</p> <table border="1"> <tr> <td colspan="2">INSPECTING OFFICER</td> </tr> <tr> <td>SIGNATURE <i>[Signature]</i></td> <td>PRINT FULL NAME RUPP, GREGORY</td> </tr> <tr> <td>TYPE II PERMIT NUMBER 230076</td> <td>TELEPHONE NUMBER (636) 537-3000</td> </tr> <tr> <td>EXPIRATION DATE 04/14/2025</td> <td></td> </tr> </table> <p>RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, Missouri Department of Health and Senior Services, by mail, fax, or e-mail</p>			<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK	<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK	<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK	<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK	<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK	<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST	<input checked="" type="checkbox"/> ETH CHECK		<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE	<input checked="" type="checkbox"/> STANDARD SUPPLIER	INTOXIMETERS		LOT# AG223501		EXP. DATE 08/23/2024	<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN		SIM. NIST EXP DATE	<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	TEST 1 @ 0.079 g/210L	TEST 2 @ 0.079 g/210L	TEST 3 @ 0.079 g/210L	REFUSALS	1	0-.04	11	.05-.09	0	.10-.14	1	.15-.19	1	OVER .19	1	INSPECTING OFFICER		SIGNATURE <i>[Signature]</i>	PRINT FULL NAME RUPP, GREGORY	TYPE II PERMIT NUMBER 230076	TELEPHONE NUMBER (636) 537-3000	EXPIRATION DATE 04/14/2025	
<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK																																																					
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK																																																					
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK																																																					
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK																																																					
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK																																																					
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST																																																					
<input checked="" type="checkbox"/> ETH CHECK																																																						
<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE																																																					
<input checked="" type="checkbox"/> STANDARD SUPPLIER	INTOXIMETERS																																																					
	LOT# AG223501																																																					
	EXP. DATE 08/23/2024																																																					
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN																																																					
	SIM. NIST EXP DATE																																																					
<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE																																																						
<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE																																																						
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE																																																						
TEST 1 @ 0.079 g/210L	TEST 2 @ 0.079 g/210L	TEST 3 @ 0.079 g/210L																																																				
REFUSALS	1	0-.04	11	.05-.09	0	.10-.14	1	.15-.19	1	OVER .19	1																																											
INSPECTING OFFICER																																																						
SIGNATURE <i>[Signature]</i>	PRINT FULL NAME RUPP, GREGORY																																																					
TYPE II PERMIT NUMBER 230076	TELEPHONE NUMBER (636) 537-3000																																																					
EXPIRATION DATE 04/14/2025																																																						



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 24-Aug-2022

Lot # AG223501 **Model** 108

Exp Date	Cyl. Type	Component	Certified Concentration
23-Aug-2024	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:08.24.2022 19:06

Approved for Release: 
Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
GREGORY RUPP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/14/2023

NUMBER 230076

EXPIRES 4/14/2025

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RUPP, GREGORY
Permit No 230076
Date Issued 4/14/2023 **Date Expires** 4/14/2025

