



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**INTOX EC/IR II MAINTENANCE REPORT**

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12820	NAME OF AGENCY WASHINGTON POLICE DEPT	DATE OF INSPECTION 03/07/2024
----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 301 Jefferson St. Washington	TIME OF INSPECTION 08:13 CST
--	---------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> <b>DIAGNOSTIC RECORD</b>	
<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST
<input checked="" type="checkbox"/> ETH CHECK	

<b>BREATH ANALYZER ACCURACY STANDARDS</b>	
<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER	INTOXIMETER LOT# AG305102 EXP. DATE 02/20/2025
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN SIM. NIST EXP DATE

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 = 0.099 g/210L	TEST 2 = 0.099 g/210L	TEST 3 = 0.099 g/210L
-----------------------	-----------------------	-----------------------

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS	0	0-.04	0	.05-.09	0	.10-.14	0	.15-.19	0	OVER .19	2
----------	---	-------	---	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME TOLLISON, DOUGLAS
TYPE II PERMIT NUMBER 230055	EXPIRATION DATE 03/27/2025
	TELEPHONE NUMBER ( 636 ) 390-1050

**RETURN COMPLETED REPORT TO THE:**  
Breath Alcohol Program, Missouri Department of Health and Senior Services,  
by mail, fax, or e-mail

STATE OF MISSOURI     )  
  )  
COUNTY OF FRANKLIN    )     SS

AFFIDAVIT

*Before me, the undersigned authority, personally appeared, Douglas A. Tollison who, being by me duly sworn, deposed as follows:*


My name is Douglas A. Tollison I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of maintenance records of the INTOX EC/IR II, SERIAL #12820. Attached hereto are 3 pages of records from the Washington Police Department for the month of March, 2024. These 3 pages of records are kept by Washington Police Department in the regular course of business, and it was the regular course of business of Washington Police Department for an employee or representative of Washington Police Department with knowledge of the act, event, condition, opinion or diagnosis recorded to make the record or to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion or diagnosis. The records attached hereto are the original or exact duplicate of the original.

Sgt. D. Tollison #285

*In witness whereof I have hereunto subscribed my name and affixed my official seal this 7 day of March, 2024.*

Stephanie L. Hellebusch  
Notary Public





STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**DOUGLAS A. TOLLISON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/27/2023

NUMBER 230055

EXPIRES 3/27/2025

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** TOLLISON, DOUGLAS  
**Permit No** 230055  
**Date Issued** 3/27/2023 **Date Expires** 3/27/2025

