



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 7:06 am, Mar 01, 2024

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER: 80-005854
LOCATION OF INSTRUMENT: LIBERTY POLICE DEPT.
DATE OF INSPECTION: 03/01/2024
TIME OF INSPECTION: 05:12

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	05:14
Cal Check	0.102	05:14
Air Blank	0.000	05:15
Cal Check	0.102	05:15
Air Blank	0.000	05:16
Cal Check	0.102	05:16
Air Blank	0.000	05:17

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE: DRY	STANDARD LOT #: AG320002	STANDARD EXPIRATION DATE: 07/19/2025
SIM TEMPERATURE: N/A	SIM SERIAL NUMBER: N/A	SIM CERTIFICATE EXPIRATION: N/A
STANDARD VALUE: 0.100	STANDARD SUPPLIER: INTOXIMETERS	
CALIBRATION CHECK RESULT 1: 0.102		
CALIBRATION CHECK RESULT 2: 0.102		
CALIBRATION CHECK RESULT 3: 0.102		
MAXIMUM DEVIATION (MUST BE WITHIN 5%): 2.0%	SPREAD (MUST BE .005 OR LESS): 0.000	

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	05:17
Subject Test	RFI*	05:17
Air Blank	0.000	05:18
*RFI Detect		

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	3	0	1	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
FUNCTIONING NORMALLY

INSPECTING OFFICER
SIGNATURE: *[Signature]* 239
PRINT NAME: BENJAMIN LAUGHLIN
TYPE II PERMIT NUMBER: 230106
EXPIRATION DATE: 05/30/2025
TELEPHONE NUMBER: 816.439.4701