



Missouri Department of Health and Senior Services  
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466  
**Margaret T. Donnelly**  
 Director



**Jeremiah W. (Jay) Nixon**  
 Governor

**Missouri Department of Health and Senior Services Breath Alcohol Program**

***CERTIFICATE OF CALIBRATION***

This report shall document that the instrument described herein was examined and tested against NIST-traceable reference standards in accordance with Missouri Department of Health and Senior Services State Public Health Laboratory Breath Alcohol Program's Testing Requirements for the Recalibration of Simulators.

**AGENCY:** Missouri State Highway Patrol, Troop E

**ADDRESS:** 4947 Highway 67 North

**CITY:** Poplar Bluff, MO

**ZIP:** 63901

**SIMULATOR INFORMATION**

**Serial Number:** 1245

**Manufacturer:** RepCo

**Model Number:** 3402C

**CALIBRATION RESULTS**

<u>Reference Temperature</u>	<u>Simulator Temperature</u>	<u>Tolerance</u>	<u>Bias</u>	<u>In Tolerance</u>	<u>Combined Uncertainty</u>
33.94	34.00	±0.1 °C	0.06 °C	YES	±0.04 °C

This calibration was performed with NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian Lutmer

This calibration was performed: 02/20/2013