Missouri Department of Health and Senior Services
P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
Margaret T. Donnelly
Director

Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Missouri State Highway Patrol
Serial Number: G11045
Manufacturer: Guth
Model Number: 34C

CALIBRATION RESULTS

Reference Temperature	Simulator Temperature	Bias	In Tolerance
33.99	34.0	-0.01 °C	YES

This calibration was performed with NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 02/18/2014

COPY OF CALIBRATION STICKER

This simulator has been calibrated according to DHSS specifications

SIMULATOR SERIAL NO.: G11045
EXPIRATION DATE: 02/18/2015
DATE OF CALIBRATION: 02/18/2014
NIST REF. THERM. SERIAL NO: 304447
BIAS: -0.01 C
ANALYST INITIALS: BML