



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Margaret T. Donnelly
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Jackson Police Department
 Serial Number: SD2272
 Manufacturer: Guth
 Model Number: 10-4D

CALIBRATION RESULTS

<u>Reference</u>	<u>Simulator</u>	<u>Bias</u>	<u>In Tolerance</u>
<u>Temperature</u>	<u>Temperature</u>		
33.98	34.0	-0.02 °C	YES

This calibration was performed with
 NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 03/07/2014

COPY OF CALIBRATION STICKER

This simulator has been calibrated according to DHSS specifications

SIMULATOR SERIAL NO.: SD2272

EXPIRATION DATE: 03/07/2015

DATE OF CALIBRATION: 03/07/2014

NIST REF. THERM SERIAL NO.: 304447

AVERAGE SIM. TEMP: 33.98 C

ANALYST INITIALS: BML