



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Gail Vasterling
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Missouri State Highway Patrol
 Serial Number: MP2129
 Manufacturer: Guth
 Model Number: 12V500

CALIBRATION RESULTS

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
34.00	34.04

This calibration was performed with
 NIST-Traceable Thermometer SN: 307715

This simulator was tested by: JLC

This testing was performed: 07/29/15

This certification expires: 07/29/16

Signature of certifying DHSS Scientist: 

Name of certifying DHSS Scientist: Brian M. Lutmer



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

BREATH ALCOHOL SIMULATOR TEST WORKSHEET

Test Simulator Information

Agency

Missouri State Highway Patrol

Email for COC

jimmy.cleveland@MissourDps.Mo.Gov

Serial Number:

MP2129

Manufacturer:

GUTH

Model Number:

12V500

NIST-Traceable Reference Thermometer Information

Serial Number:

307715

Date of Certification:

08/19/14

Date of Expiration:

08/19/15

Test Simulator Measurements

Readings	Reference Thermometer	Test Simulator
1	34.00	34.04
2	34.00	34.04
3	34.00	34.04
4	34.00	34.04
5	34.00	34.04

Bias (δ_T):

+04

Technician performing testing:

Jimmy L. Cleveland

I hereby certify that all data submitted within this form was collected in accordance with the DHSS Procedure for the Testing of Breath Alcohol Simulators and 19 CSR 25-30.051, Breath Analyzer Calibration and Accuracy Verification Standards.

Signature:

Jimmy L. Cleveland

Date:

7-29-15

Submit completed forms for simulator certification to DHSS Breath Alcohol Program by fax at (573) 840-9139 or by email at <http://www.health.mo.gov> or brea.alcohol@health.mo.gov.

COPY OF LABEL PLACED ON SIMULATOR

Please simulator has been collected according to DHSS procedure.

SIMULATOR SERIAL NO.: MP2129
 EXPIRATION DATE: 7-29-15
 DATE OF CALIBRATION: 8-19-14
 NIST TRACABLE SERIAL NO.: 307715
 AVERAGE SD: 0.04
 NAME: JLC