



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Gail Vasterling
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: MISSOURI STATE HIGHWAY PATROL
 Serial Number: MP2454
 Manufacturer: Guth
 Model Number: 12V500

CALIBRATION RESULTS

<u>Reference</u>	<u>Simulator</u>
<u>Temperature</u>	<u>Temperature</u>
34.01	33.98

This calibration was performed with
 NIST-Traceable Thermometer SN: 306168

This simulator was tested by: JLC

This testing was performed: 10/19/2015

This certification expires: 10/19/2016

Signature of certifying DHSS
 Scientist: 

Name of certifying DHSS Scientist: Ellen Strawsine

