Missouri Department of Health and Senior Services
P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
Gail Vasterling
Director

Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: **Jackson Police Department**

Serial Number: **SD2272**

Manufacturer: **Guth**

Model Number: **10-4D**

CALIBRATION RESULTS

Reference Temperature	Simulator Temperature
33.98	34.0

This calibration was performed with NIST-Traceable Thermometer SN: 304454

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 03/12/2015

COPY OF CALIBRATION STICKER

This simulator has been calibrated according to DHSS specifications

SIMULATOR SERIAL NO.: SD2272
EXPIRATION DATE: 03/12/2016
DATE OF CALIBRATION: 03/12/2015
NIST REF. THEM. SERIAL NO.: 304454
AVERAGE SIM. TEMP: 33.98 C
ANALYST INITIALS: BML