



Missouri Department of Health and Senior Services  
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466  
 Margaret T. Donnelly  
 Director



Jeremiah W. (Jay) Nixon  
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

***SIMULATOR CALIBRATION REPORT***

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

**SIMULATOR INFORMATION**

Agency: Miner Police Department  
 Serial Number: SD2273  
 Manufacturer: Guth  
 Model Number: 10-4D

**CALIBRATION RESULTS**

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
33.97	34.0

This calibration was performed with  
 NIST-Traceable Thermometer SN: 304454

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 01/13/2015

**COPY OF CALIBRATION STICKER**

This simulator has been calibrated according to DHSS specifications

	SIMULATOR SERIAL NO.:	<u>SD2273</u>
	EXPIRATION DATE:	<u>01/13/2016</u>
	DATE OF CALIBRATION:	<u>01/13/2015</u>
	NIST REF. THEM. SERIAL NO.:	<u>304454</u>
	AVERAGE SIM. TEMP:	<u>33.97 C</u>
	ANALYST INITIALS:	<u>BML</u>