



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19 CSR 25-30.051 (4).



Fulton
Police Department

Avg. Temp. 33.96
Min. Temp. 33.96
Max. Temp. 33.97

Therm. Serial # 093752

CHECKED BY: DRL

Serial # **SD 2726**

Expiration Date: 04/15/2016



Date Checked: 04/15/2015

Technician Printed Name: DAN LUCAS

Technician Signature: [Signature]

Date: 4/15/15

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834