



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Margaret T. Donnelly
 Director



Jeremiah W. (Jay) Nixon
 Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: Scott City Police Department
 Serial Number: SD2730
 Manufacturer: Guth
 Model Number: 10-4D

CALIBRATION RESULTS

| <u>Reference Temperature</u> | <u>Simulator Temperature</u> |
|------------------------------|------------------------------|
| 33.94 | 34.0 |

This calibration was performed with NIST-Traceable Thermometer SN: 304454

This calibration was performed by: Brian M. Lutmer

This calibration was performed: 01/21/2015

COPY OF CALIBRATION STICKER

This simulator has been calibrated according to DHSS specifications



SIMULATOR SERIAL NO.: SD2730
 EXPIRATION DATE: 01/21/2016
 DATE OF CALIBRATION: 01/21/2015
 NIST REF. THEM. SERIAL NO.: 304454
 AVERAGE SIM. TEMP.: 33.94 C
 ANALYST INITIALS: BML